

LB 114 223


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34296



A= 0.0000e+000

B= 2.8586e-004

C= 7.8042e-003

Rho= 0.9993272

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.0	0.000	0.015	0.015	24	0.000	24					-
0.2	0.200	0.195	-0.005	656	0.0 %	656					-3
2.5	2.500	2.398	-0.102	8360	0.0 %	8360					-4
5.0	5.000	5.030	0.030	17569	0.0 %	17569					1
7.5	7.500	7.751	0.251	27086	0.0 %	27086					3
10.0	10.000	9.812	-0.188	34296	0.0 %	34296					-2

LB1214223 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Apr 2021 11:32:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	24	-	0.0000	7470A	PPB	30 Apr 2021 11:35:43	S	Std
0.2 - 1	50.2	656	-	0.2000	7470A	PPB	30 Apr 2021 11:41:55	S	Std
2.5 - 1	52.5	8360	-	2.5000	7470A	PPB	30 Apr 2021 11:44:11	S	Std
5.0 - 1	55	17569	-	5.0000	7470A	PPB	30 Apr 2021 11:46:28	S	Std
7.5 - 1	57.5	27086	-	7.5000	7470A	PPB	30 Apr 2021 11:48:44	S	Std
10.0 - 1	59	34296	-	10.0000	7470A	PPB	30 Apr 2021 11:53:35	S	Std
ICV11 - 1	ICV11	13468	3.8578	-	7470A	PPB	30 Apr 2021 11:56:31	U	SMPL
ICB11 - 1	ICB11	-204	-0.0505	-	7470A	PPB	30 Apr 2021 11:58:46	U	SMPL
CCV24 - 1	CCV24	18125	5.1890	-	7470A	PPB	30 Apr 2021 12:01:01	U	SMPL
CCB24 - 1	CCB24	-195	-0.0479	-	7470A	PPB	30 Apr 2021 12:03:16	U	SMPL
CRA - 1	CRA	621	0.1853	-	7470A	PPB	30 Apr 2021 12:05:32	U	SMPL
HighStd - 1	HighStd	34093	9.7537	-	7470A	PPB	30 Apr 2021 12:07:47	U	SMPL
ChkStd - 1	ChkStd	24775	7.0900	-	7470A	PPB	30 Apr 2021 12:10:03	U	SMPL
PB136068BL - 1	PBW	-274	-0.0705	-	7470A	PPB	30 Apr 2021 12:12:20	U	SMPL
PB136068BS - 1	LCSW	14085	4.0342	-	7470A	PPB	30 Apr 2021 12:14:37	U	SMPL
M2202-02 - 1	290	-174	-0.0419	-	7470A	PPB	30 Apr 2021 12:16:52	U	SMPL
M2203-01 - 1	309	-32	-0.0013	-	7470A	PPB	30 Apr 2021 12:19:09	U	SMPL
M2204-02 - 1	OILY-DEBRIS-1907-1910	-5	0.0064	-	7470A	PPB	30 Apr 2021 12:21:26	U	SMPL
M2222-02 - 1	TP-5-COMP	13	0.0115	-	7470A	PPB	30 Apr 2021 12:23:42	U	SMPL
M2222-07 - 1	TP-5A-COMP	40	0.0192	-	7470A	PPB	30 Apr 2021 12:25:58	U	SMPL
CCV25 - 1	CCV25	18105	5.1833	-	7470A	PPB	30 Apr 2021 12:28:13	U	SMPL
CCB25 - 1	CCB25	-156	-0.0368	-	7470A	PPB	30 Apr 2021 12:30:28	U	SMPL
M2222-07DUP - 1	TP-5A-COMPDUP	-1	0.0075	-	7470A	PPB	30 Apr 2021 12:32:44	U	SMPL
M2222-07MS - 1	TP-5A-COMPMS	13279	3.8037	-	7470A	PPB	30 Apr 2021 12:34:59	U	SMPL
M2222-07MSD - 1	TP-5A-COMPMSD	12704	3.6394	-	7470A	PPB	30 Apr 2021 12:37:14	U	SMPL
PB136068TB - 1	PB136053TB	-146	-0.0339	-	7470A	PPB	30 Apr 2021 12:39:30	U	SMPL
M2222-07LX5 - 1		0	0.0078	-	7470A	PPB	30 Apr 2021 12:41:45	U	SMPL
M2222-07A - 1		13276	3.8029	-	7470A	PPB	30 Apr 2021 12:44:01	U	SMPL
CCV26 - 1	CCV26	18109	5.1845	-	7470A	PPB	30 Apr 2021 12:46:17	U	SMPL
CCB26 - 1	CCB26	-171	-0.0411	-	7470A	PPB	30 Apr 2021 12:48:35	U	SMPL