



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114248

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	05/03/2021	09:25
2	M2202-08		1	8	100.00	>212.0	>212.0	05/03/2021	09:45
3	M2202-08DUP		1	8	100.00	>212.0	>212.0	05/03/2021	10:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp5321 WorkList ID : 148525 Department : Wet-Chemistry Date : 05-03-2021 09:02:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/05/2021	Water	M2202-08	C	Cool 4 deg C	PSEG03	O41	TOTE-COMP	04/28/2021	1010A

Date/Time 05/03/21 09:05
 Raw Sample Received by: FS (Welcher)
 Raw Sample Relinquished by: RJ (ELT-126)

Date/Time 05/03/21 10:15
 Raw Sample Received by: RJ (ELT-126)
 Raw Sample Relinquished by: FS (Welcher)