



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114249

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	05/03/2021	09:25
2	M2204-03		1	7	95.00	203.0	203.3	05/03/2021	10:45
3	M2204-03DUP		1	7	95.00	203.0	203.3	05/03/2021	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp5321-1

WorkList ID : 148526

Department : Wet-Chemistry

Date : 05-03-2021 09:03:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/05/2021	Solid	M2204-03	Flash Point	Cool 4 deg C	PSEG03	O41	DIESEL-1900	04/28/2021	1010A

Date/Time

05/03/21 10:15

Raw Sample Received by:

KS (Weber)

Raw Sample Relinquished by:

RJ (EXT-196)

Date/Time

05/03/21 11:25

Raw Sample Received by:

RJ (EXT-196)

Raw Sample Relinquished by:

KS (Weber)