



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 5/4/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 05/03/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:18
Out Date: 05/04/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB114250

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M2253-01	COMP-41772-75	23	1.12	9.88	7.14	68.7	
M2253-02	COMP-41772-75	24	1.16	9.79	9.00	90.8	
M2253-03	31069	25	1.12	9.5	9.09	95.1	
M2253-04	31069	26	1.17	9.69	9.26	95.0	
M2255-01	OK-01-050321	27	1.18	9.82	9.36	94.7	
M2255-02	OK-01-050321-EPH-2	28	1.17	9.68	9.24	94.8	
M2256-01	R1-A-D-COMP	1	1.15	9.66	8.79	89.8	
M2256-03	R1-A-BOTTOM	2	1.19	9.71	8.97	91.3	
M2256-12	R2-A-D-COMP	3	1.14	9.75	8.58	86.4	
M2256-14	R2-A-BOTTOM	4	1.16	9.86	8.75	87.2	
M2258-01	N096290-10	5	1.15	9.94	1.49	3.9	SLUDGE SAMPLE
M2258-02	N096423-01	6	1.12	9.64	1.37	2.9	SLUDGE SAMPLE
M2258-03	N096405-05	7	1.15	9.97	1.58	4.9	SLUDGE SAMPLE
M2258-04	N096405-06	8	1.18	9.98	1.28	1.1	SLUDGE SAMPLE
M2258-05	N096405-07	9	1.16	9.76	1.83	7.8	SLUDGE SAMPLE
M2258-06	N096405-08	10	1.15	9.97	1.7	6.2	SLUDGE SAMPLE
M2258-07	N096326-09	11	1.15	9.98	1.78	7.1	SLUDGE SAMPLE
M2259-01	JEC658L-1-1	15	1.00	2.00	2.00	100.0	PILC REEL
M2259-02	JEC658L-1-2	16	1.00	2.00	2.00	100.0	PILC REEL
M2259-03	BCZ58713-1-1	17	1.00	2.00	2.00	100.0	PILC REEL
M2259-04	BCZ58713-1-2	18	1.00	2.00	2.00	100.0	PILC REEL
M2259-05	BCZ83039-1-1	19	1.00	2.00	2.00	100.0	PILC REEL
M2259-06	BCZ83039-1-2	20	1.00	2.00	2.00	100.0	PILC REEL
M2259-07	BC258873-1-1	21	1.00	2.00	2.00	100.0	PILC REEL
M2259-08	BC258873-1-2	22	1.00	2.00	2.00	100.0	PILC REEL
M2263-01	S-1	12	1.14	9.63	8.83	90.6	
M2263-01D	S-1DUP	13	1.15	9.65	8.84	90.5	
M2263-02	S-1A	14	1.19	9.74	9.00	91.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$