



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114251

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	05/03/2021	11:50
2	M2214-06		1	6	54.00	129.2	129.5	05/03/2021	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : M2214 Wet PT WorkList ID : 148426 Department : Wet-Chemistry Date : 04-28-2021 11:56:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/11/2021	Solid	M2214-01	Anions Group1	Cool 4 deg C	CHEM02	QA Office	PT-AN-SOIL	04/26/2021	9056A
04/26/2021	Solid	M2214-02	Corrosivity	Cool 4 deg C	CHEM02	QA Office	PT-CORR-SOIL	04/26/2021	9045D
04/26/2021	Solid	M2214-03	pH	Cool 4 deg C	CHEM02	QA Office	PT-CORR-SOIL	04/26/2021	9041A
05/10/2021	Solid	M2214-04	Cyanide	Cool 4 deg C	CHEM02	QA Office	PT-CN-SOIL	04/26/2021	9012B
05/10/2021	Solid	M2214-05	Cyanide	Cool 4 deg C	CHEM02	QA Office	PT-CN-SOIL	04/26/2021	9014
05/10/2021	Solid	M2214-06	Flash Point	Cool 4 deg C	CHEM02	QA Office	PT-FP-SOIL	04/26/2021	1010A
05/11/2021	Solid	M2214-07	Hexavalent Chromium	Cool 4 deg C	CHEM02	QA Office	PT-CR6-SOIL	04/26/2021	7196A
05/11/2021	Solid	M2214-08	Ammonia	Cool 4 deg C	CHEM02	QA Office	PT-NUT-SOIL	04/26/2021	SM4500-NH3
05/11/2021	Solid	M2214-08	Phosphorus, Total	Cool 4 deg C	CHEM02	QA Office	PT-NUT-SOIL	04/26/2021	365.3
05/11/2021	Solid	M2214-08	TKN	Cool 4 deg C	CHEM02	QA Office	PT-NUT-SOIL	04/26/2021	SM4500 N Org
05/11/2021	Solid	M2214-08	TOC	Cool 4 deg C	CHEM02	QA Office	PT-NUT-SOIL	04/26/2021	9060A
05/11/2021	Solid	M2214-09	TOC	Cool 4 deg C	CHEM02	QA Office	PT-NUT-SOIL	04/26/2021	Lloyd Kahn
05/11/2021	Solid	M2214-10	Oil and Grease	Cool 4 deg C	CHEM02	QA Office	PT-OGR-SOIL	04/26/2021	9071B
05/10/2021	Solid	M2214-23	TS	Cool 4 deg C	CHEM02	QA Office	PT-SOL-SOIL	04/26/2021	SM2540 B

Date/Time 4/28/21 12:05 Date/Time 5/11/21 12:30
 Raw Sample Received by: AS (wetchem) Raw Sample Received by: AS (wetchem)
 Raw Sample Relinquished by: MA (Cibola Murray) Raw Sample Relinquished by: AS (wetchem)