


LB114254

7471B

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.0040e-004

C= -3.1655e-002

Rho= 0.9998774

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.0	0.000	-0.022	-0.022	47	0.000	47					0.0
0.2	0.200	0.176	-0.024	1034	0.0 %	1034					-12
2.5	2.500	2.483	-0.017	12546	0.0 %	12546					-1
5.0	5.000	5.124	0.124	25727	0.0 %	25727					2
7.5	7.500	7.497	-0.003	37571	0.0 %	37571					6
10.0	10.000	9.943	-0.057	49774	0.0 %	49774					-1

LB114254 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 May 2021 15:54:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.0000	7471B	PPB	03 May 2021 15:57:52	S	Std
0.2 - 1	50.2	1034	-	0.2000	7471B	PPB	03 May 2021 16:00:09	S	Std
2.5 - 1	52.5	12546	-	2.5000	7471B	PPB	03 May 2021 16:02:26	S	Std
5.0 - 1	55	25727	-	5.0000	7471B	PPB	03 May 2021 16:04:43	S	Std
7.5 - 1	57.5	37571	-	7.5000	7471B	PPB	03 May 2021 16:06:59	S	Std
10.0 - 1	510	49774	-	10.0000	7471B	PPB	03 May 2021 16:11:50	S	Std
ICV14 - 1	ICV14	20202	4.0167	-7471B	PPB	03 May 2021 16:14:34	U	SMPL	
ICB14 - 1	ICB14	-61	-0.0439	-7471B	PPB	03 May 2021 16:16:49	U	SMPL	
CCV34 - 1	CCV34	26092	5.1971	-7471B	PPB	03 May 2021 16:19:05	U	SMPL	
CCB34 - 1	CCB34	-84	-0.0485	-7471B	PPB	03 May 2021 16:21:20	U	SMPL	
CRA - 1	CRA	1035	0.1758	-7471B	PPB	03 May 2021 16:23:36	U	SMPL	
HighStd - 1	HighStd	50362	10.0607	-7471B	PPB	03 May 2021 16:25:51	U	SMPL	
ChkStd - 1	ChkStd	34506	6.8832	-7471B	PPB	03 May 2021 16:28:05	U	SMPL	
PB136130BL - 1	PBS	-173	-0.0663	-7471B	PPB	03 May 2021 16:30:23	U	SMPL	
PB136130BS - 1	LCSS	18583	3.6923	-7471B	PPB	03 May 2021 16:32:41	U	SMPL	
M2235-01 - 1	OR-02-043021	10117	1.9957	-7471B	PPB	03 May 2021 16:34:57	U	SMPL	
M2236-02 - 1	11991	1564	0.2818	-7471B	PPB	03 May 2021 16:37:14	U	SMPL	
M2237-01 - 1	JC-02-043021	19516	3.8793	-7471B	PPB	03 May 2021 16:39:32	U	SMPL	
M2238-01 - 1	HR-02-043021	6585	1.2880	-7471B	PPB	03 May 2021 16:41:48	U	SMPL	
M2239-01 - 1	TR-08-043021	3120	0.5936	-7471B	PPB	03 May 2021 16:44:05	U	SMPL	
CCV35 - 1	CCV35	24131	4.8041	-7471B	PPB	03 May 2021 16:46:20	U	SMPL	
CCB35 - 1	CCB35	-68	-0.0453	-7471B	PPB	03 May 2021 16:48:35	U	SMPL	
M2242-01 - 1	STOCK-PILE	4437	0.8575	-7471B	PPB	03 May 2021 16:50:51	U	SMPL	
M2242-01DUP - 1	STOCK-PILEDUP	4036	0.7771	-7471B	PPB	03 May 2021 16:53:06	U	SMPL	
M2242-01MS - 1	STOCK-PILEMS	24332	4.8444	-7471B	PPB	03 May 2021 16:55:21	U	SMPL	
M2242-01MSD - 1	STOCK-PILEMSD	23014	4.5803	-7471B	PPB	03 May 2021 16:57:37	U	SMPL	
M2242-01LX5 - 1		732	0.1150	-7471B	PPB	03 May 2021 16:59:53	U	SMPL	
M2242-01A - 1		24606	4.8993	-7471B	PPB	03 May 2021 17:02:10	U	SMPL	
CCV36 - 1	CCV36	24559	4.8899	-7471B	PPB	03 May 2021 17:04:27	U	SMPL	
CCB36 - 1	CCB36	-91	-0.0499	-7471B	PPB	03 May 2021 17:06:45	U	SMPL	