



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114262

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/04/2021	09:05
2	CAL2	1	Water	NA	NA	20.3	7.00	05/04/2021	09:06
3	CAL3	1	Water	NA	NA	20.3	10.03	05/04/2021	09:07
4	ICV	1	Water	NA	NA	20.3	7.01	05/04/2021	09:09
5	CCV1	1	Water	NA	NA	20.4	2.01	05/04/2021	09:10
6	M2253-05	1	Water	NA	NA	20.2	8.39	05/04/2021	09:11
7	M2253-05DUP	1	Water	NA	NA	20.2	8.40	05/04/2021	09:12
8	CCV2	1	Water	NA	NA	20.4	12.02	05/04/2021	09:15

2B 114262

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2253 WorkList ID : 148552 Department : Wet-Chemistry Date : 05-04-2021 08:57:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/03/2021	Water	M2253-05	pH	Cool 4 deg C	PSEG03	O61	41764	05/03/2021	9040C

Date/Time 05/04/21 9:00
Raw Sample Received by: JB (SM) (200)
Raw Sample Relinquished by: CP (SM)

Date/Time 05/04/21 10:00
Raw Sample Received by: CP (SM)
Raw Sample Relinquished by: JB (SM) (200)