



Analytical Summary Report

Analysis Method: 9045D
Parameter: Corrosivity
Run Number: LB114263
BalanceID: WC SC-4

Analyst By : JIGNESH
Supervisor Review By : apatel
Slope : 98.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/04/2021	09:30
2	CAL2	1	Water	NA	NA	20.3	7.00	05/04/2021	09:31
3	CAL3	1	Water	NA	NA	20.3	10.03	05/04/2021	09:32
4	ICV	1	Water	NA	NA	20.3	6.98	05/04/2021	09:35
5	CCV1	1	Water	NA	NA	20.3	2.01	05/04/2021	09:36
6	M2202-02	1	Solid	20.02	20	21.2	12.45	05/04/2021	09:39
7	M2252-01	1	Solid	20.03	20	21.0	5.26	05/04/2021	09:42
8	M2252-04	1	Solid	20.05	20	20.9	5.48	05/04/2021	09:44
9	M2252-04DUP	1	Solid	20.04	20	20.8	5.49	05/04/2021	09:45
10	CCV2	1	Water	NA	NA	20.4	12.02	05/04/2021	09:46

18114263

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity m2256 WorkList ID : 148539 Department : Wet-Chemistry Date : 05-03-2021 14:47:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/28/2021	Solid	M2202-02	Corrosivity	Cool 4 deg C	PSEG03	O41	290	04/28/2021	9045D
05/03/2021	Solid	M2252-01	Corrosivity	Cool 4 deg C	PSEG03	O61	STONE-1	05/03/2021	9045D
05/03/2021	Solid	M2252-04	Corrosivity	Cool 4 deg C	PSEG03	O61	STONE-2	05/03/2021	9045D

05/04/21 9:00
 Date/Time 05/04/21
 Raw Sample Received by: JP (not Chen)
 Raw Sample Relinquished by: Ch SM

05/04/21 10:35
 Date/Time 05/04/21
 Raw Sample Received by: Ch SM
 Raw Sample Relinquished by: JP (not Chen)