



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114266

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	05/04/2021	12:25
2	M2253-05		1	9	100.00	>212.0	>212.0	05/04/2021	12:45
3	M2253-05DUP		1	9	100.00	>212.0	>212.0	05/04/2021	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp5421 WorkList ID : 148559 Department : Wet-Chemistry Date : 05-04-2021 09:09:10

Due Date	Matrix	Sample	Test	Flash Point	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/10/2021	Water	M2253-05	F	Flash Point	Cool 4 deg C	PSEG03	O61	41764	05/03/2021	1010A

Date/Time 05/04/21 12:00
 Raw Sample Received by: KS (Welch)
 Raw Sample Relinquished by: RJ (Welch)

Date/Time 05/04/21 13:15
 Raw Sample Received by: RJ (Welch)
 Raw Sample Relinquished by: KS (Welch)