

LB114270


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

55254



A= 0.0000e+000

B= 1.7683e-004

C= -3.6315e-002

Rho= 0.9989185

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.027	-0.027	51	0.000	51				
0.2	0.200	0.174	-0.026	1192	0.0 %	1192				
2.5	2.500	2.394	-0.106	13741	0.0 %	13741				
5.0	5.000	5.174	0.174	29467	0.0 %	29467				
7.5	7.500	7.751	0.251	44037	0.0 %	44037				
10.0	10.000	9.734	-0.266	55254	0.0 %	55254				

4/6/11
-13
-4
3
3
-3

LB114270 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 May 2021 14:35:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	51	-	0.0000	7471B	PPB	04 May 2021 14:37:30	S	Std
0.2 - 1	50.2	1192	-	0.2000	7471B	PPB	04 May 2021 14:41:04	S	Std
2.5 - 1	52.5	13741	-	2.5000	7471B	PPB	04 May 2021 14:43:20	S	Std
5.0 - 1	55	29467	-	5.0000	7471B	PPB	04 May 2021 14:45:37	S	Std
7.5 - 1	57.5	44037	-	7.5000	7471B	PPB	04 May 2021 14:47:53	S	Std
10.0 - 1	57.5	55254	-	10.0000	7471B	PPB	04 May 2021 14:50:11	S	Std
ICV16 - 1	ICV16	23359	4.0943	-	7471B	PPB	04 May 2021 14:53:47	U	SMPL
ICB16 - 1	ICB16	-91	-0.0524	-	7471B	PPB	04 May 2021 14:56:02	U	SMPL
CCV40 - 1	CCV40	29660	5.2085	-	7471B	PPB	04 May 2021 14:58:18	U	SMPL
CCB40 - 1	CCB40	-113	-0.0563	-	7471B	PPB	04 May 2021 15:00:33	U	SMPL
CRA - 1	CRA	1158	0.1685	-	7471B	PPB	04 May 2021 15:02:51	U	SMPL
HighStd - 1	HighStd	56231	9.9070	-	7471B	PPB	04 May 2021 15:05:06	U	SMPL
ChkStd - 1	ChkStd	40809	7.1799	-	7471B	PPB	04 May 2021 15:07:21	U	SMPL
PB136150BL - 1	PBS	-226	-0.0763	-	7471B	PPB	04 May 2021 15:09:38	U	SMPL
PB136150BS - 1	LCSS	20297	3.5528	-	7471B	PPB	04 May 2021 15:14:34	U	SMPL
M2252-01 - 1	STONE-1	6494	1.1120	-	7471B	PPB	04 May 2021 15:16:49	U	SMPL
M2252-04 - 1	STONE-2	1227	0.1807	-	7471B	PPB	04 May 2021 15:19:06	U	SMPL
M2253-01 - 1	COMP-41772-75	1538	0.2356	-	7471B	PPB	04 May 2021 15:21:23	U	SMPL
M2253-03 - 1	31069	1877	0.2956	-	7471B	PPB	04 May 2021 15:23:40	U	SMPL
M2254-01 - 1	TP-1	14381	2.5067	-	7471B	PPB	04 May 2021 15:27:58	U	SMPL
CCV41 - 1	CCV41	28871	5.0689	-	7471B	PPB	04 May 2021 15:30:13	U	SMPL
CCB41 - 1	CCB41	-64	-0.0476	-	7471B	PPB	04 May 2021 15:32:27	U	SMPL
M2254-05 - 1	TP-2	16865	2.9459	-	7471B	PPB	04 May 2021 15:34:44	U	SMPL
M2255-01 - 1	OK-01-050321	1904	0.3004	-	7471B	PPB	04 May 2021 15:36:59	U	SMPL
M2256-01 - 1	R1-A-D-COMP	1390	0.2095	-	7471B	PPB	04 May 2021 15:39:14	U	SMPL
M2256-12 - 1	R2-A-D-COMP	2377	0.3840	-	7471B	PPB	04 May 2021 15:41:29	U	SMPL
M2263-01 - 1	S-1	1141	0.1654	-	7471B	PPB	04 May 2021 15:43:45	U	SMPL
M2263-01DUP - 1	S-1DUP	1111	0.1601	-	7471B	PPB	04 May 2021 15:46:01	U	SMPL
M2263-01MS - 1	S-1MS	19612	3.4317	-	7471B	PPB	04 May 2021 15:48:17	U	SMPL
M2263-01MSD - 1	S-1MSD	20142	3.5254	-	7471B	PPB	04 May 2021 15:50:34	U	SMPL
M2263-01LX5 - 1		229	0.0042	-	7471B	PPB	04 May 2021 15:52:51	U	SMPL
M2263-01A - 1		23169	4.0607	-	7471B	PPB	04 May 2021 15:55:06	U	SMPL
CCV42 - 1	CCV42	29576	5.1936	-	7471B	PPB	04 May 2021 15:57:22	U	SMPL
CCB42 - 1	CCB42	-21	-0.0400	-	7471B	PPB	04 May 2021 15:59:38	U	SMPL