



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114275

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/04/2021	17:00
2	CAL2	1	Water	NA	NA	20.2	7.00	05/04/2021	17:01
3	CAL3	1	Water	NA	NA	20.2	10.03	05/04/2021	17:02
4	ICV	1	Water	NA	NA	20.3	6.99	05/04/2021	17:05
5	CCV1	1	Water	NA	NA	20.3	2.01	05/04/2021	17:06
6	M2266-01	1	Solid	20.03	20	20.8	8.52	05/04/2021	17:10
7	M2266-01DUP	1	Solid	20.04	20	20.9	8.53	05/04/2021	17:11
8	M2269-01	1	Solid	20.02	20	20.3	5.56	05/04/2021	17:12
9	M2275-01	1	Solid	20.03	20	20.8	7.90	05/04/2021	17:15
10	M2275-09	1	Solid	20.04	20	20.7	7.48	05/04/2021	17:16
11	CCV2	1	Water	NA	NA	20.4	12.02	05/04/2021	17:19

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity m2269 WorkList ID : 148573 Department : Wet-Chemistry Date : 05-04-2021 15:15:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/04/2021	Solid	M2266-01	Corrosivity	Cool 4 deg C	PSEG03		CRT-016	05/04/2021	9045D
05/03/2021	Solid	M2269-01	Corrosivity	Cool 4 deg C	PSEG03	O63	1153	05/03/2021	9045D
05/04/2021	Solid	M2275-01	Corrosivity	Cool 4 deg C	PSEG03	O63	CC-STONES	05/04/2021	9045D
05/04/2021	Solid	M2275-09	Corrosivity	Cool 4 deg C	PSEG03	O63	CRT-020	05/04/2021	9045D

Date/Time 05/04/21 15:45
 Raw Sample Received by: JO Not Empty
 Raw Sample Relinquished by: KS (Wetchem)

Date/Time 05/04/21 17:35
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: JO Not Empty