



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114281

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.00	05/05/2021	09:20
2	CAL2	1	Water	NA	NA	20.1	7.00	05/05/2021	09:21
3	CAL3	1	Water	NA	NA	20.2	10.02	05/05/2021	09:22
4	ICV	1	Water	NA	NA	20.3	7.00	05/05/2021	09:24
5	CCV1	1	Water	NA	NA	20.3	2.01	05/05/2021	09:25
6	M2292-01	1	Water	NA	NA	20.3	6.82	05/05/2021	09:26
7	M2292-01DUP	1	Water	NA	NA	20.4	6.83	05/05/2021	09:27
8	CCV2	1	Water	NA	NA	20.3	12.02	05/05/2021	09:30

28 114281

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2292 WorkList ID : 148595 Department : Wet-Chemistry Date : 05-05-2021 09:12:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/04/2021	Water	M2292-01	pH	Cool 4 deg C	APT101	P61	WGGM-21-WWW-TSF-019	05/04/2021	SM 4500-H B

Date/Time 05/05/21 9:15
Raw Sample Received by: JP (104 Chem)
Raw Sample Relinquished by: JP (104 Chem)

Date/Time 05/05/21 10:15
Raw Sample Received by: JP (104 Chem)
Raw Sample Relinquished by: JP (104 Chem)