



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB114285

Analyst By : JIGNESH
Supervisor Review By : apatel
Slope : 98.7
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/05/2021	11:10
2	CAL2	1	Water	NA	NA	20.2	7.00	05/05/2021	11:11
3	CAL3	1	Water	NA	NA	20.2	10.03	05/05/2021	11:12
4	ICV	1	Water	NA	NA	20.2	6.99	05/05/2021	11:15
5	CCV1	1	Water	NA	NA	20.3	2.01	05/05/2021	11:16
6	M2298-01	1	Water	NA	NA	20.2	6.92	05/05/2021	11:19
7	M2298-01DUP	1	Water	NA	NA	20.3	6.93	05/05/2021	11:20
8	CCV2	1	Water	NA	NA	20.3	12.02	05/05/2021	11:22

2B 114285

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity m2298 WorkList ID : 148606 Department : Wet-Chemistry Date : 05-05-2021 10:51:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/04/2021	Water	M2298-01	P Corrosivity	Cool 4 deg C	TETR06	P53	CAL-AQIDW-20210504	05/04/2021	9040C

Date/Time 05/05/21 11:00
 Raw Sample Received by: JP (wet Chem)
 Raw Sample Relinquished by: JP (SM)

Date/Time 05/05/21 12:00
 Raw Sample Received by: JP (SM)
 Raw Sample Relinquished by: JP (wet Chem)