

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 05/07/2021

OVEN IN TEMP1: 104 °C

OVEN OUT TEMP1: 104 °C

DATE & TIME IN1: 05/06/2021 16:00

DATE & TIME OUT1: 05/07/2021 08:00

OVEN IN TEMP2: 103 °C

OVEN OUT TEMP2: 103 °C

DATE & TIME IN2: 05/07/2021 08:30

DATE & TIME OUT2: 05/07/2021 10:10

Run Number: LB114304

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB114304BL	LB114304BL	148.5671	148.5671	100	148.5672	148.5672	148.5672	0.0001	1
2	M2307-03	MW-05	142.7694	142.7694	100	143.6410	143.6410	143.6410	0.8716	8716
3	M2307-04	MW-18	152.1601	152.1601	100	152.6283	152.6283	152.6283	0.4682	4682
4	M2307-04DUP	MW-18DUP	152.9691	152.9691	100	153.4461	153.4461	153.4461	0.4770	4770

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

18114304

WORKLIST(Hardcopy Internal Chain)

WorkList Name: TS M2307 WorkList ID: 148643 Department: Wet-Chemistry Date: 05-06-2021 08:38:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/12/2021	Water	M2307-03	TS	Cool 4 deg C	LIRO01	P51	MW-05	05/05/2021	SM2540 B
05/12/2021	Water	M2307-04	TS	Cool 4 deg C	LIRO01	P51	MW-18	05/05/2021	SM2540 B

Date/Time 05/06/21 12:10
 Raw Sample Received by: JR Joff (5ul)
 Raw Sample Relinquished by: JR Joff (5ul)

Date/Time 05/06/21 16:10
 Raw Sample Received by: JR Joff (5ul)
 Raw Sample Relinquished by: JR Joff (5ul)