



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114308

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	05/06/2021	09:25
2	M2269-01		1	7	100.00	>212.0	>212.0	05/06/2021	09:50
3	M2269-01DUP		1	7	100.00	>212.0	>212.0	05/06/2021	10:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 114308

WorkList Name : fp5521 WorkList ID : 148605 Department : Wet-Chemistry Date : 05-05-2021 10:19:40

Due Date	Matrix	Sample	Test	Flash Point	Cool 4 deg C	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/05/2021	Solid	M2269-01	B	Flash Point	Cool 4 deg C	PSEG03	O63	1153	05/03/2021	1010A

Date/Time 05/06/21 09:05
 Raw Sample Received by: KS (Weberhen)
 Raw Sample Relinquished by: PJ C Ext - (Cub)

Date/Time 05/06/21 10:30
 Raw Sample Received by: RJ C Ext - (Cub)
 Raw Sample Relinquished by: KS (Weberhen)