



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114311

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/06/2021	13:15
2	CAL2	1	Water	NA	NA	20.3	7.00	05/06/2021	13:16
3	CAL3	1	Water	NA	NA	20.4	10.03	05/06/2021	13:19
4	ICV	1	Water	NA	NA	20.4	6.99	05/06/2021	13:20
5	CCV1	1	Water	NA	NA	20.3	2.01	05/06/2021	13:22
6	M2295-09	1	Water	NA	NA	20.3	13.01	05/06/2021	13:25
7	M2295-10	1	Water	NA	NA	20.4	13.02	05/06/2021	13:26
8	M2295-11	1	Water	NA	NA	20.3	12.93	05/06/2021	13:28
9	M2322-11	1	Water	NA	NA	20.7	7.13	05/06/2021	13:30
10	M2322-11DUP	1	Water	NA	NA	20.8	7.14	05/06/2021	13:31
11	CCV2	1	Water	NA	NA	20.4	12.02	05/06/2021	13:32

WORKLIST(Hardcopy Internal Chain)

VB 11431)

WorkList Name : CORROSIONITY M2322 WorkList ID : 148661 Department : Wet-Chemistry Date : 05-06-2021 13:07:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/05/2021	Water	M2295-09	Corrosivity	Cool 4 deg C	PSEG03	O63	DRUM-B	05/05/2021	9040C
05/05/2021	Water	M2295-10	Corrosivity	Cool 4 deg C	PSEG03	O63	DRUM-C	05/05/2021	9040C
05/05/2021	Water	M2295-11	Corrosivity	Cool 4 deg C	PSEG03	O63	DRUM-E	05/05/2021	9040C
05/05/2021	Water	M2322-11	Corrosivity	Cool 4 deg C	PARS02	P51	WC-20210505	05/05/2021	9040C

Date/Time 05/06/21 13:16
 Raw Sample Received by: JP (wt (mem)
 Raw Sample Relinquished by: JP (sm)

Date/Time 05/06/21 14:20
 Raw Sample Received by: JP (sm)
 Raw Sample Relinquished by: JP (wt (mem)