



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114318

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	05/06/2021	11:50
2	M2295-09		1	9	100.00	>212.0	>212.0	05/06/2021	12:12
3	M2295-09DUP		1	9	100.00	>212.0	>212.0	05/06/2021	12:40
4	M2295-10		1	8	100.00	>212.0	>212.0	05/06/2021	13:05
5	M2295-11		1	8	100.00	>212.0	>212.0	05/06/2021	13:35
6	M2307-03		1	7	100.00	>212.0	>212.0	05/06/2021	14:10
7	M2307-04		1	8	100.00	>212.0	>212.0	05/06/2021	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP5621

WorkList ID : 148657

Department : Wet-Chemistry

Date : 05-06-2021 11:07:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/12/2021	Water	M2295-09	Flash Point	Cool 4 deg C	PSEG03	O63	DRUM-B	05/05/2021	1010A
05/12/2021	Water	M2295-10	Flash Point	Cool 4 deg C	PSEG03	O63	DRUM-C	05/05/2021	1010A
05/12/2021	Water	M2295-11	Flash Point	Cool 4 deg C	PSEG03	O63	DRUM-E	05/05/2021	1010A
05/12/2021	Water	M2307-03	Flash Point	Cool 4 deg C	LIRO01	P51	MW-05	05/05/2021	1010A
05/12/2021	Water	M2307-04	Flash Point	Cool 4 deg C	LIRO01	P51	MW-18	05/05/2021	1010A

Date/Time 05/06/21 11:30
 Raw Sample Received by: KS (Weber)
 Raw Sample Relinquished by: RJ (EXT-100)

Date/Time 05/06/21 14:45
 Raw Sample Received by: RJ (EXT-100)
 Raw Sample Relinquished by: KS (Weber)