



TOTAL VOLATILE SOLIDS 160.4

TEMP1 IN: 104 °C 05/06/2021 16:00 TEMP1 OUT: 104 °C 05/07/2021 08:00
TEMP2 IN: 103 °C 05/07/2021 08:30 TEMP2 OUT: 103 °C 05/07/2021 10:00
TEMP3 IN: 540 °C 05/07/2021 10:30 TEMP3 OUT: 540 °C 05/07/2021 12:00
TEMP4 IN: 560 °C 05/07/2021 12:30 TEMP4 OUT: 560 °C 05/07/2021 14:00

Run Number: LB114322

SUPERVISOR: apatel

ANALYST: JIGNESH

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Empty Dish + Sample Weight (g)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (%)
1	LB114322BL	62.8743	62.8743	83.5271	62.8744	62.8744	62.8744	62.8744	0.0000	0
2	M2323-01	59.5621	59.5621	79.6016	76.6824	76.6824	75.8696	75.8696	0.8128	4.7
3	M2323-05	58.3181	58.3181	79.2487	73.2160	73.216	72.1026	72.1026	1.1134	7.5
4	M2323-09	60.8013	60.8013	81.2249	79.2490	79.249	78.8387	78.8387	0.4103	2.2
5	M2323-11	55.8883	55.8883	75.8339	73.7537	73.7537	73.3844	73.3844	0.3693	2.1
6	M2323-13	59.9807	59.9807	79.7043	77.7282	77.7282	77.4101	77.4101	0.3181	1.8
7	M2323-15	59.3956	59.3956	79.9366	77.7398	77.7398	77.3203	77.3203	0.4195	2.3
8	M2323-17	53.5526	53.5526	74.1847	72.4292	72.4292	72.2374	72.2374	0.1918	1
9	M2323-17DUP	60.9421	60.9421	81.4778	79.7934	79.7934	79.6104	79.6104	0.1830	1

A = Sample Weight (g)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{F - E} \star 100$

VB114322

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tvs m2323 WorkList ID : 148670 Department : Wet-Chemistry Date : 05-06-2021 15:00:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/11/2021	Solid	M2323-01	TVS	Cool 4 deg C	ENTA05	P61	AST2-COPR3-1	05/05/2021	160.4
05/11/2021	Solid	M2323-05	TVS	Cool 4 deg C	ENTA05	P61	AST2-COPR2-1	05/05/2021	160.4
05/11/2021	Solid	M2323-09	TVS	Cool 4 deg C	ENTA05	P61	AST2-WC1-A	05/04/2021	160.4
05/11/2021	Solid	M2323-11	TVS	Cool 4 deg C	ENTA05	P61	AST2-WC1-C	05/04/2021	160.4
05/11/2021	Solid	M2323-13	TVS	Cool 4 deg C	ENTA05	P61	AST2-WC1-D	05/04/2021	160.4
05/11/2021	Solid	M2323-15	TVS	Cool 4 deg C	ENTA05	P61	AST2-WC1-E	05/04/2021	160.4
05/11/2021	Solid	M2323-17	TVS	Cool 4 deg C	ENTA05	P61	AST2-WC3-A	05/04/2021	160.4

Date/Time 05/06/21 15:00
Raw Sample Received by: JP 1004 (m)
Raw Sample Relinquished by: OP SMI

Date/Time 05/06/21 17:35
Raw Sample Received by: OP SMI
Raw Sample Relinquished by: JP 1004 (m)