



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114323

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	05/06/2021	17:10
2	CAL2	1	Water	NA	NA	20.4	7.01	05/06/2021	17:11
3	CAL3	1	Water	NA	NA	20.4	10.03	05/06/2021	17:12
4	ICV	1	Water	NA	NA	20.3	7.00	05/06/2021	17:15
5	CCV1	1	Water	NA	NA	20.3	2.01	05/06/2021	17:16
6	M2214-02	1	Solid	20.02	20	20.5	6.97	05/06/2021	17:19
7	M2295-03	1	Solid	20.03	20	20.7	12.49	05/06/2021	17:22
8	M2295-07	1	Solid	20.02	20	20.6	12.14	05/06/2021	17:24
9	M2323-02	1	Solid	20.03	20	20.8	10.89	05/06/2021	17:26
10	M2323-06	1	Solid	20.04	20	20.3	9.16	05/06/2021	17:29
11	M2323-06DUP	1	Solid	20.03	20	20.4	9.17	05/06/2021	17:30
12	CCV2	1	Water	NA	NA	20.4	12.02	05/06/2021	17:31

WORKLIST(Hardcopy Internal Chain)

LB 114323

WorkList Name : corrosivity m2323

WorkList ID : 148672

Department : Wet-Chemistry

Date : 05-06-2021 15:12:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/05/2021	Solid	M2295-03	Corrosivity	Cool 4 deg C	PSEG03	O63	DRUMS-A-B-C	05/05/2021	9045D
05/05/2021	Solid	M2295-07	Corrosivity	Cool 4 deg C	PSEG03	O63	DRUMS-D-E	05/05/2021	9045D
05/05/2021	Solid	M2323-02	Corrosivity	Cool 4 deg C	ENTA05	P61	AST2-COPR3-1	05/05/2021	9045D
05/05/2021	Solid	M2323-06	Corrosivity	Cool 4 deg C	ENTA05	P61	AST2-COPR2-1	05/05/2021	9045D

Date/Time 05/06/21 15:12:00
 Raw Sample Received by: JO (wet) [Signature]
 Raw Sample Relinquished by: CS/SMI

Date/Time 05/06/21 17:14:00
 Raw Sample Received by: CS/SMI
 Raw Sample Relinquished by: JO (wet) [Signature]