



Analytical Summary Report

Analysis Method: 9041A

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114324

Slope : NA

BalanceID: WC SC-4

pH Meter ID : NA

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807
3.0-5.5 pH paper refill	W1934
Test Paper, PH, Short Range 6.0/8.0	W2350
Test Paper,PH Short Range 9.0/10.0	W1939
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.0	05/06/2021	17:10
2	CAL2	1	Water	NA	NA	20.4	7.0	05/06/2021	17:11
3	CAL3	1	Water	NA	NA	20.4	10.0	05/06/2021	17:12
4	ICV	1	Water	NA	NA	20.3	7.0	05/06/2021	17:15
5	CCV1	1	Water	NA	NA	20.3	2.0	05/06/2021	17:16
6	M2214-03	1	Solid	20.02	20	20.3	7.0	05/06/2021	17:19
7	CCV2	1	Water	NA	NA	20.4	12.0	05/06/2021	17:31