



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114334

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/07/2021	12:45
2	CAL2	1	Water	NA	NA	20.3	7.00	05/07/2021	12:46
3	CAL3	1	Water	NA	NA	20.3	10.03	05/07/2021	12:47
4	ICV	1	Water	NA	NA	20.3	6.99	05/07/2021	12:48
5	CCV1	1	Water	NA	NA	20.2	2.01	05/07/2021	12:50
6	M2323-01	1	Solid	20.03	20	20.8	10.89	05/07/2021	13:00
7	M2323-05	1	Solid	20.04	20	20.3	9.16	05/07/2021	13:03
8	M2323-09	1	Solid	20.02	20	20.5	9.60	05/07/2021	13:00
9	M2323-11	1	Solid	20.01	20	20.6	9.27	05/07/2021	13:07
10	M2323-13	1	Solid	20.03	20	20.5	9.06	05/07/2021	13:10
11	M2323-15	1	Solid	20.04	20	20.4	10.03	05/07/2021	13:11
12	M2323-17	1	Solid	20.02	20	20.4	9.53	05/07/2021	13:12
13	M2334-01	1	Solid	20.03	20	20.4	8.83	05/07/2021	13:15
14	M2334-01DUP	1	Solid	20.04	20	20.3	8.84	05/07/2021	13:16
15	CCV2	1	Water	NA	NA	20.4	12.02	05/07/2021	13:19

13114333

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2334 WorkList ID : 148704 Department : Wet-Chemistry Date : 05-07-2021 12:33:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/05/2021	Solid	M2323-01	pH	Cool 4 deg C	ENTA05	P61	AST2-COPR3-1	05/05/2021	9045D
05/05/2021	Solid	M2323-05	pH	Cool 4 deg C	ENTA05	P61	AST2-COPR2-1	05/05/2021	9045D
05/04/2021	Solid	M2323-09	pH	Cool 4 deg C	ENTA05	P61	AST2-WC1-A	05/04/2021	9045D
05/04/2021	Solid	M2323-11	pH	Cool 4 deg C	ENTA05	P61	AST2-WC1-C	05/04/2021	9045D
05/04/2021	Solid	M2323-13	pH	Cool 4 deg C	ENTA05	P61	AST2-WC1-D	05/04/2021	9045D
05/04/2021	Solid	M2323-15	pH	Cool 4 deg C	ENTA05	P61	AST2-WC1-E	05/04/2021	9045D
05/04/2021	Solid	M2323-17	pH	Cool 4 deg C	ENTA05	P61	AST2-WC3-A	05/04/2021	9045D
05/07/2021	Solid	M2334-01	pH	Cool 4 deg C	PSEG03	O63	CHRT-20111	05/07/2021	9045D

Date/Time 05/07/21 12:33
 Raw Sample Received by: JP Wet Chem
 Raw Sample Relinquished by: CP/SM/

Date/Time 05/07/21 12:00
 Raw Sample Received by: CP/SM/
 Raw Sample Relinquished by: JP Wet Chem