



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 5/10/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 05/07/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 10:05
Out Date: 05/08/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB114344

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M2333-08	BP-VPB-RW8-GW-658-660	1	1.15	9.98	1.59	5.0	SLUDGE SAMPLE
M2334-01	CHRT-20111	2	1.16	9.94	8.58	84.5	
M2335-01	OK-02-050721	3	1.15	9.9	8.47	83.7	
M2335-02	OK-02-050721-EPH-2	4	1.14	9.94	9.03	89.7	
M2335-02D	OK-02-050721-EPH-2DUP	5	1.16	9.96	9.02	89.3	
M2341-05	SVOC-GPC-BLANK	6	1.00	2.00	2.00	100.0	
M2341-06	PEST-GPC-BLANK	7	1.00	2.00	2.00	100.0	
M2341-07	PEST-GPC-BLANK-SPIKE	8	1.00	2.00	2.00	100.0	
M2341-08	PCB-GPC-BLANK	9	1.00	2.00	2.00	100.0	
M2341-09	PCB-GPC-BLANK-SPIKE	10	1.00	2.00	2.00	100.0	
M2341-10	SVOC-GCP2-BLANK	11	1.00	2.00	2.00	100.0	
M2341-11	PEST-GCP2-BLANK	12	1.00	2.00	2.00	100.0	
M2341-12	PEST-GCP2-BLANK-SPIKE	13	1.00	2.00	2.00	100.0	
M2341-13	PCB-GCP2-BLANK	14	1.00	2.00	2.00	100.0	
M2341-14	PCB-GCP2-BLANK-SPIKE	15	1.00	2.00	2.00	100.0	
M2343-01	C-1-1	16	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2343-02	C-1-2	17	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2343-03	C-1-3	18	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2343-04	C-2-1	19	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2343-05	C-2-2	20	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2343-06	C-2-3	21	1.00	2.00	2.00	100.0	CONCRETE SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$