



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114360

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/10/2021	11:00
2	CAL2	1	Water	NA	NA	20.2	7.00	05/10/2021	11:01
3	CAL3	1	Water	NA	NA	20.2	10.03	05/10/2021	11:02
4	ICV	1	Water	NA	NA	20.3	6.99	05/10/2021	11:05
5	CCV1	1	Water	NA	NA	20.2	2.01	05/10/2021	11:06
6	M2345-01	1	Water	NA	NA	20.1	6.26	05/10/2021	11:10
7	M2345-01DUP	1	Water	NA	NA	20.2	6.27	05/10/2021	11:11
8	CCV2	1	Water	NA	NA	20.3	12.02	05/10/2021	11:12

2B 114360

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m2345 WorkList ID : 148740 Department : Wet-Chemistry Date : 05-10-2021 10:37:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/05/2021	Water	M2345-01	pH	Cool 4 deg C	STER04	D11	POOL-WATER	05/05/2021	9040C

Date/Time 05.10.21 10:40
Raw Sample Received by: JB (Wet Chem)
Raw Sample Relinquished by: CP SM

Date/Time 05.10.21 11:40
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JB (Wet Chem)