



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114368

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	05/10/2021	10:25
2	M2322-11		1	7	100.00	>212.0	>212.0	05/10/2021	10:50
3	M2322-11DUP		1	7	100.00	>212.0	>212.0	05/10/2021	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114368

WorkList Name : fp5721 WorkList ID : 148692 Department : Wet-Chemistry Date : 05-07-2021 09:12:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/18/2021	Water	M2322-11	Flash Point	Cool 4 deg C	PARS02	P51	WC-20210505	05/05/2021	1010A

Date/Time 05/07/21 10:00
 Raw Sample Received by: KS (wekher)
 Raw Sample Relinquished by: RC CEXT 1ab

Date/Time 05/07/21 11:25
 Raw Sample Received by: RC CEXT 1ab
 Raw Sample Relinquished by: KS (wekher)