

LB114375

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092


Reviewed by : AM Instrument ID : Konelab
5/11/2021 9:42

Test CNEPA-NEW

Accepted 5/11/2021 9:41
~~Factor (slope)~~ ~~1039 0.000463~~ 0.000963
Bias (y-intercept) 0.008
05/11/2021 AM

Coeff. of det. 0.999975

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors %RE
1	0.0PPBCN	0.007	-0.4841	0.0000	-12.0
2	5.0PPBCN	0.012	4.4021	5.0000	-6.7
3	10PPBCN	0.017	9.3838	10.0000	1.3
4	100PPBCN	0.105	101.3283	100.0000	0.5
5	250PPBCN	0.250	251.2345	250.0000	-0.2
6	500PPBCN	0.489	499.1353	500.0000	

5/11/2021 AM

LB114375

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092Reviewed by : AM Instrument ID : Konelab

5/11/2021 10:27

Test: CNEPA-NEW

43/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV001 ICV001	94.141	0.0	0.099	
ICB001 ICB001	0.362	0.0	0.008	
CCV001 CCV001	241.013	0.0	0.240	
CCB001 CCB001	0.121	0.0	0.008	
P B136282BL PBW282	0.014	0.0	0.008	
M250-4-MDL - WATER-QT3-2021-05	4.630	0.0	0.012	
P B138283 PBS283	0.323	0.0	0.008	
M250-03MDL - SOIL-QT3-2021-05	5.427	0.0	0.013	
CCV002 CCV002	244.018	0.0	0.243	
CCB002 CCB002	0.257	0.0	0.008	

N 10
Mean 59.031
SD 100.9398
CV% 171.00

Aquakem v. 7.2AQ1

Results from time period:

Tue May 11 09:18:02 2021

Tue May 11 10:23:10 2021

Sample Id	Sam	Test shc	Test t	Result	Result unit	Result date and time	Stat
S0.0	A	CNEPA-IP		-0.4841	µg/l	5/11/2021 9:36:35	
S5.0	A	CNEPA-IP		4.4021	µg/l	5/11/2021 9:36:36	
S10.0	A	CNEPA-IP		9.3838	µg/l	5/11/2021 9:36:37	
S100.0	A	CNEPA-IP		101.3283	µg/l	5/11/2021 9:36:38	
S250.0	A	CNEPA-IP		251.2345	µg/l	5/11/2021 9:36:39	
S500.0	A	CNEPA-IP		499.1353	µg/l	5/11/2021 9:36:40	
ICV001 ICV001	S	CNEPA-IP		94.1412	µg/l	5/11/2021 10:16:23	
ICB001 ICB001	S	CNEPA-IP		0.3616	µg/l	5/11/2021 10:16:26	
CCV001 CCV001	S	CNEPA-IP		241.0129	µg/l	5/11/2021 10:16:27	
CCB001 CCB001	S	CNEPA-IP		0.1205	µg/l	5/11/2021 10:16:30	
PB136282BL PBW282	S	CNEPA-IP		0.0143	µg/l	5/11/2021 10:16:32	
M2250-01 MDL-WATER-QT3-2021-05	S	CNEPA-IP		4.6297	µg/l	5/11/2021 10:23:02	
PB138283 PBS283	S	CNEPA-IP		0.3234	µg/l	5/11/2021 10:23:03	
M2250-03 MDL-SOIL-QT3-2021-05	S	CNEPA-IP		5.4273	µg/l	5/11/2021 10:23:06	
CCV002 CCV002	S	CNEPA-IP		244.0181	µg/l	5/11/2021 10:23:08	
CCB002 CCB002	S	CNEPA-IP		0.2574	µg/l	5/11/2021 10:23:09	