

LB114376

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : Am Instrument ID : Konelab

5/11/2021 11:29

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.381	0.0	0.075	
ICB1	0.139	0.0	0.001	
CCV1	242.377	0.0	0.191	
CCB1	-0.044	0.0	0.000	
PB136308BL	0.133	0.0	0.001	
M2334-01	0.088	0.0	0.001	
M2334-01DUP	0.209	0.0	0.001	
CCV2	249.250	0.0	0.196	
CCB2	0.190	0.0	0.001	

N	9
Mean	65.303
SD	106.9990
CV%	163.85

Aquakem v. 7.2AQ1

Results from time period:

Tue May 11 10:55:20 2021

Tue May 11 11:26:22 2021

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
0.0PPBCN	A	Reactive CN	P	-0.2009	µg/l	5/11/2021 10:55:20
5.0PPBCN	A	Reactive CN	P	4.6916	µg/l	5/11/2021 10:55:21
10PPBCN	A	Reactive CN	P	9.4412	µg/l	5/11/2021 10:55:22
50PPBCN	A	Reactive CN	P	49.2769	µg/l	5/11/2021 10:55:23
100PPBCN	A	Reactive CN	P	100.8563	µg/l	5/11/2021 10:55:24
250PPBCN	A	Reactive CN	P	252.0392	µg/l	5/11/2021 10:55:25
500PPBCN	A	Reactive CN	P	498.8957	µg/l	5/11/2021 10:55:26
ICV1	S	Reactive CN	P	95.3811	µg/l	5/11/2021 11:20:33
ICB1	S	Reactive CN	P	0.1388	µg/l	5/11/2021 11:20:35
CCV1	S	Reactive CN	P	242.3771	µg/l	5/11/2021 11:20:37
CCB1	S	Reactive CN	P	-0.0437	µg/l	5/11/2021 11:20:40
PB136308BL	S	Reactive CN	P	0.1329	µg/l	5/11/2021 11:20:42
M2334-01	S	Reactive CN	P	0.0878	µg/l	5/11/2021 11:20:43
M2334-01DUP	S	Reactive CN	P	0.209	µg/l	5/11/2021 11:26:18
CCV2	S	Reactive CN	P	249.2505	µg/l	5/11/2021 11:26:19
CCB2	S	Reactive CN	P	0.1897	µg/l	5/11/2021 11:26:21

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

5/11/2021 10:57

Reviewed by : AM

Instrument ID : Konelab

Test Total CN


Accepted 5/11/2021 10:57

Factor 1274

Bias 0

Coeff. of det. 0.999966

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors %RE
1	0.0PPBCN	0.000	-0.2009	0.0000	-6.2
2	5.0PPBCN	0.004	4.6916	5.0000	-5.6
3	10PPBCN	0.008	9.4412	10.0000	-5.4
4	50PPBCN	0.039	49.2769	50.0000	0.9
5	100PPBCN	0.080	100.8563	100.0000	0.8
6	250PPBCN	0.198	252.0392	250.0000	-0.2
7	500PPBCN	0.392	498.8957	500.0000	

05/11/2021
AM