

LB114379


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs.:

39478



A= 0.0000e+000

B= 2.5025e-004

C= -1.3014e-003

Rho= 0.9996918

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	0.008	0.008	38	0.000	38					1
0.2	0.200	0.192	-0.008	771	0.0 %	771					-4
2.5	2.500	2.458	-0.042	9829	0.0 %	9829					-2
5.0	5.000	4.981	-0.019	19910	0.0 %	19910					0
7.5	7.500	7.682	0.182	30704	0.0 %	30704					2
10.0	10.000	9.878	-0.122	39478	0.0 %	39478					-1

LB114379 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 May 2021 13:24:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	38	-	0.0000	7470A	PPB	11 May 2021 13:29:02	S	Std
0.2 - 1	50.2	771	-	0.2000	7470A	PPB	11 May 2021 13:31:19	S	Std
2.5 - 1	52.5	9829	-	2.5000	7470A	PPB	11 May 2021 13:33:35	S	Std
5.0 - 1	55	19910	-	5.0000	7470A	PPB	11 May 2021 13:35:53	S	Std
7.5 - 1	57.5	30704	-	7.5000	7470A	PPB	11 May 2021 13:38:08	S	Std
10.0 - 1	5/b	39478	-	10.0000	7470A	PPB	11 May 2021 13:43:08	S	Std
ICV18 - 1	ICV18	15418	3.8571	-	7470A	PPB	11 May 2021 13:45:48	U	SMPL
ICB18 - 1	ICB18	-130	-0.0338	-	7470A	PPB	11 May 2021 13:48:02	U	SMPL
CCV46 - 1	CCV46	20860	5.2189	-	7470A	PPB	11 May 2021 13:50:18	U	SMPL
CCB46 - 1	CCB46	-109	-0.0286	-	7470A	PPB	11 May 2021 13:52:33	U	SMPL
CRA - 1	CRA	753	0.1871	-	7470A	PPB	11 May 2021 13:54:49	U	SMPL
HighStd - 1	HighStd	39629	9.9159	-	7470A	PPB	11 May 2021 13:57:03	U	SMPL
ChkStd - 1	ChkStd	28649	7.1681	-	7470A	PPB	11 May 2021 13:59:18	U	SMPL
PB136236BL - 1	PBW	-11	-0.0041	-	7470A	PPB	11 May 2021 14:05:06	U	SMPL
PB136236BS - 1	LCSW	15550	3.8901	-	7470A	PPB	11 May 2021 14:10:10	U	SMPL
M2252-02 - 1	STONE-1	-99	-0.0261	-	7470A	PPB	11 May 2021 14:12:26	U	SMPL
M2252-05 - 1	STONE-2	-3	-0.0021	-	7470A	PPB	11 May 2021 14:14:43	U	SMPL
M2254-02 - 1	TP-1	1	-0.0011	-	7470A	PPB	11 May 2021 14:17:00	U	SMPL
M2254-02DUP - 1	TP-1DUP	2	-0.0008	-	7470A	PPB	11 May 2021 14:19:22	U	SMPL
M2254-02MS - 1	TP-1MS	15529	3.8849	-	7470A	PPB	11 May 2021 14:24:17	U	SMPL
CCV47 - 1	CCV47	20886	5.2254	-	7470A	PPB	11 May 2021 14:26:32	U	SMPL
CCB47 - 1	CCB47	-121	-0.0316	-	7470A	PPB	11 May 2021 14:28:48	U	SMPL
M2254-02MSD - 1	TP-1MSD	15463	3.8683	-	7470A	PPB	11 May 2021 14:31:04	U	SMPL
M2254-06 - 1	TP-2	-104	-0.0273	-	7470A	PPB	11 May 2021 14:33:20	U	SMPL
M2256-02 - 1	R1-A-D-COMP	-2	-0.0018	-	7470A	PPB	11 May 2021 14:35:36	U	SMPL
M2256-13 - 1	R2-A-D-COMP	24	0.0047	-	7470A	PPB	11 May 2021 14:37:51	U	SMPL
M2265-02 - 1	124055	46	0.0102	-	7470A	PPB	11 May 2021 14:40:07	U	SMPL
M2269-01 - 1	1153	78	0.0182	-	7470A	PPB	11 May 2021 14:42:23	U	SMPL
PB136318BL - 1	PBW	123	0.0295	-	7470A	PPB	11 May 2021 14:44:39	U	SMPL
PB136318BS - 1	LCSW	13900	3.4772	-	7470A	PPB	11 May 2021 14:49:29	U	SMPL
M2295-03 - 1	DRUM-A-B-C	-12	-0.0043	-	7470A	PPB	11 May 2021 14:51:46	U	SMPL
M2295-03DUP - 1	DRUM-A-B-CDUP	111	0.0265	-	7470A	PPB	11 May 2021 14:54:02	U	SMPL
CCV48 - 1	CCV48	18806	4.7049	-	7470A	PPB	11 May 2021 14:56:18	U	SMPL
CCB48 - 1	CCB48	-134	-0.0348	-	7470A	PPB	11 May 2021 14:58:34	U	SMPL
M2295-03MS - 1	DRUM-A-B-CMS	13700	3.4271	-	7470A	PPB	11 May 2021 15:00:51	U	SMPL
M2295-03MSD - 1	DRUM-A-B-CMSD	13270	3.3195	-	7470A	PPB	11 May 2021 15:03:07	U	SMPL
M2295-07 - 1	DRUM-D-E	-20	-0.0063	-	7470A	PPB	11 May 2021 15:05:23	U	SMPL
M2323-02 - 1	AST2-COPR3-1	103	0.0245	-	7470A	PPB	11 May 2021 15:07:41	U	SMPL
M2323-06 - 1	AST2-COPR2-1	177	0.0430	-	7470A	PPB	11 May 2021 15:09:56	U	SMPL
PB136236TB - 1	PB136138TB	79	0.0185	-	7470A	PPB	11 May 2021 15:12:12	U	SMPL
PB136318TB - 1	PB136215TB	92	0.0217	-	7470A	PPB	11 May 2021 15:14:28	U	SMPL
CCV49 - 1	CCV49	18042	4.5137	-	7470A	PPB	11 May 2021 15:16:45	U	SMPL
CCB49 - 1	CCB49	-110	-0.0288	-	7470A	PPB	11 May 2021 15:19:02	U	SMPL
M2254-02LX5 - 1		2	-0.0008	-	7470A	PPB	11 May 2021 15:21:19	U	SMPL
M2254-02A - 1		14341	3.5876	-	7470A	PPB	11 May 2021 15:23:35	U	SMPL
M2295-03LX5 - 1		-68	-0.0183	-	7470A	PPB	11 May 2021 15:25:50	U	SMPL
M2295-03A - 1		14900	3.7274	-	7470A	PPB	11 May 2021 15:28:07	U	SMPL
CCV50 - 1	CCV50	20258	5.0683	-	7470A	PPB	11 May 2021 15:30:22	U	SMPL
CCB50 - 1	CCB50	-133	-0.0346	-	7470A	PPB	11 May 2021 15:32:39	U	SMPL