

LB114386


7471B

INSTRUMENT ID: CVL

Linear

μ Abs.:

45674



A= 0.0000e+000

B= 2.1795e-004

C= -9.3093e-002

Rho= 0.9996849

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9013
0.0	0.000	-0.024	-0.024	317	0.000	317					2
0.2	0.200	0.204	0.004	1365	0.0 %	1365					-2
2.5	2.500	2.443	-0.057	11635	0.0 %	11635					1
5.0	5.000	5.056	0.056	23624	0.0 %	23624					2
7.5	7.500	7.659	0.159	35569	0.0 %	35569					-1
10.0	10.000	9.862	-0.138	45674	0.0 %	45674					

LB114386 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 11 May 2021 15:41:50					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	317	-	0.0000	7471B	PPB	11 May 2021 15:43:02	S	Std
0.2 - 1	SP	1365	-	0.2000	7471B	PPB	11 May 2021 15:45:19	S	Std
2.5 - 1	SP	11635	-	2.5000	7471B	PPB	11 May 2021 15:47:37	S	Std
5.0 - 1	SP	23624	-	5.0000	7471B	PPB	11 May 2021 15:49:56	S	Std
7.5 - 1	SP	35569	-	7.5000	7471B	PPB	11 May 2021 15:52:14	S	Std
10.0 - 1	SP	45674	-	10.0000	7471B	PPB	11 May 2021 15:54:32	S	Std
ICV19 - 1	ICV19	19290	4.1112	-	7471B	PPB	11 May 2021 15:57:16	U	SMPL
ICB19 - 1	ICB19	239	-0.0410	-	7471B	PPB	11 May 2021 16:02:04	U	SMPL
CCV51 - 1	CCV51	23860	5.1073	-	7471B	PPB	11 May 2021 16:04:19	U	SMPL
CCB51 - 1	CCB51	393	-0.0074	-	7471B	PPB	11 May 2021 16:09:08	U	SMPL
CRA - 1	CRA	1393	0.2105	-	7471B	PPB	11 May 2021 16:11:28	U	SMPL
HighStd - 1	HighStd	45628	9.8517	-	7471B	PPB	11 May 2021 16:13:43	U	SMPL
ChkStd - 1	ChkStd	31892	6.8579	-	7471B	PPB	11 May 2021 16:15:58	U	SMPL
PB136320BL - 1	PBS	149	-0.0606	-	7471B	PPB	11 May 2021 16:18:16	U	SMPL
PB136320BS - 1	LCSS	15740	3.3375	-	7471B	PPB	11 May 2021 16:20:34	U	SMPL
M2265-01 - 1	124055	78852	17.0929	-	7471B	PPB	11 May 2021 16:22:50	U	SMPL
M2266-01 - 1	CRT-016	2642	0.4827	-	7471B	PPB	11 May 2021 16:25:08	U	SMPL
M2275-01 - 1	CC-STONES	203	-0.0488	-	7471B	PPB	11 May 2021 16:27:30	U	SMPL
M2275-09 - 1	CRT-020	2446	0.4400	-	7471B	PPB	11 May 2021 16:29:46	U	SMPL
M2278-01 - 1	SU-01-050421	4371	0.8596	-	7471B	PPB	11 May 2021 16:32:01	U	SMPL
CCV52 - 1	CCV52	23961	5.1293	-	7471B	PPB	11 May 2021 16:34:16	U	SMPL
CCB52 - 1	CCB52	123	-0.0663	-	7471B	PPB	11 May 2021 16:36:31	U	SMPL
M2279-01 - 1	EME-12	4496	0.8868	-	7471B	PPB	11 May 2021 16:43:41	U	SMPL
M2279-01DUP - 1	EME-12DUP	4600	0.9095	-	7471B	PPB	11 May 2021 16:45:56	U	SMPL
M2279-01MS - 1	EME-12MS	25843	5.5395	-	7471B	PPB	11 May 2021 16:50:44	U	SMPL
M2279-01MSD - 1	EME-12MSD	25715	5.5116	-	7471B	PPB	11 May 2021 16:52:59	U	SMPL
M2279-02 - 1	EME-13	2756	0.5076	-	7471B	PPB	11 May 2021 16:55:17	U	SMPL
M2279-03 - 1	EME-14	1473	0.2280	-	7471B	PPB	11 May 2021 16:57:35	U	SMPL
M2279-04 - 1	EME-15	1471	0.2275	-	7471B	PPB	11 May 2021 16:59:51	U	SMPL
M2279-05 - 1	EME-16	1504	0.2347	-	7471B	PPB	11 May 2021 17:02:08	U	SMPL
M2279-06 - 1	EME-17	1509	0.2358	-	7471B	PPB	11 May 2021 17:04:25	U	SMPL
M2279-07 - 1	EME-18	1926	0.3267	-	7471B	PPB	11 May 2021 17:06:40	U	SMPL
CCV53 - 1	CCV53	25203	5.4000	-	7471B	PPB	11 May 2021 17:08:56	U	SMPL
CCB53 - 1	CCB53	144	-0.0617	-	7471B	PPB	11 May 2021 17:11:10	U	SMPL
M2279-08 - 1	EME-19	3062	0.5743	-	7471B	PPB	11 May 2021 17:13:27	U	SMPL
M2280-01 - 1	FK-PS-01-072	721	0.0641	-	7471B	PPB	11 May 2021 17:15:42	U	SMPL
M2309-01 - 1	MH-TP1	3398	0.6475	-	7471B	PPB	11 May 2021 17:17:58	U	SMPL
M2309-03 - 1	MH-TP2	3058	0.5734	-	7471B	PPB	11 May 2021 17:20:14	U	SMPL
M2310-01 - 1	BS-SP1	1257	0.1809	-	7471B	PPB	11 May 2021 17:22:30	U	SMPL
M2310-03 - 1	BS-SP2	1814	0.3023	-	7471B	PPB	11 May 2021 17:24:46	U	SMPL
M2279-01LX5 - 1		1211	0.1708	-	7471B	PPB	11 May 2021 17:27:03	U	SMPL
M2279-01A - 1		24579	5.2640	-	7471B	PPB	11 May 2021 17:29:20	U	SMPL
M2265-01DLX5 - 1		16784	3.5650	-	7471B	PPB	11 May 2021 17:31:37	U	SMPL
CCV54 - 1	CCV54	24915	5.3372	-	7471B	PPB	11 May 2021 17:33:54	U	SMPL
CCB54 - 1	CCB54	154	-0.0595	-	7471B	PPB	11 May 2021 17:36:10	U	SMPL