

LBW4388

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AM

Instrument ID : Konelab

5/11/2021 15:08

Test: Phenolics-

Sample Id	Result ^{mg/L}	Dil. 1 +	Response	Errors
ICV1	0.955	0.0	0.106	
ICB1	-0.006	0.0	0.004	
CCV1	0.971	0.0	0.108	
CCB1	-0.005	0.0	0.004	
PB136273BL	-0.004	0.0	0.004	
PB136273BS	0.962	0.0	0.107	
M2262-07	0.023	0.0	0.007	
M2262-08	0.047	0.0	0.010	
PB136274BL	0.000	0.0	0.005	
PB136274BS	0.963	0.0	0.107	
M2262-01	0.021	0.0	0.007	
M2262-02	0.045	0.0	0.009	
CCV2	0.959	0.0	0.107	
CCB2	-0.004	0.0	0.004	

N	14
Mean	0.352
SD	0.4722
CV%	134.17

Aquakem v. 7.2AQ1

Results from time period:

Tue May 11 14:50:11 2021

Tue May 11 15:05:10 2021

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Phenolics-	P	-0.0004	mg/l	5/11/2021 13:30:16	
0.05PPM	A	Phenolics-	P	0.0509	mg/l	5/11/2021 13:30:17	
0.1PPM	A	Phenolics-	P	0.1004	mg/l	5/11/2021 13:30:18	
0.25PPM	A	Phenolics-	P	0.2532	mg/l	5/11/2021 13:30:19	
0.50PPM	A	Phenolics-	P	0.4978	mg/l	5/11/2021 13:30:20	
1.0PPM	A	Phenolics-	P	0.996	mg/l	5/11/2021 13:30:21	
2.0PPM	A	Phenolics-	P	2.0021	mg/l	5/11/2021 13:30:22	
ICV1	S	Phenolics-	P	0.9552	mg/l	5/11/2021 14:50:11	
ICB1	S	Phenolics-	P	-0.0058	mg/l	5/11/2021 14:50:14	
CCV1	S	Phenolics-	P	0.9715	mg/l	5/11/2021 14:50:15	
CCB1	S	Phenolics-	P	-0.0052	mg/l	5/11/2021 14:50:17	
PB136273BL	S	Phenolics-	P	-0.0042	mg/l	5/11/2021 14:50:20	
PB136273BS	S	Phenolics-	P	0.9617	mg/l	5/11/2021 14:50:21	
M2262-07	S	Phenolics-	P	0.0227	mg/l	5/11/2021 15:00:13	
M2262-08	S	Phenolics-	P	0.047	mg/l	5/11/2021 15:00:15	
PB136274BL	S	Phenolics-	P	0.0003	mg/l	5/11/2021 15:00:18	
PB136274BS	S	Phenolics-	P	0.9627	mg/l	5/11/2021 15:00:20	
M2262-01	S	Phenolics-	P	0.0214	mg/l	5/11/2021 15:00:21	
M2262-02	S	Phenolics-	P	0.0448	mg/l	5/11/2021 15:00:23	
CCV2	S	Phenolics-	P	0.9588	mg/l	5/11/2021 15:05:07	
CCB2	S	Phenolics-	P	-0.0038	mg/l	5/11/2021 15:05:09	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

5/11/2021 13:32

Reviewed by : AM Instrument ID : Konelab


Test Phenolics-

Accepted 5/11/2021 13:32

Factor 9.375
Bias 0.005

Coeff. of det. 0.999988

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors	%RE
1	0.00PPM	0.005	-0.0004	0.0000		
2	Phenol-2	0.010	0.0509	0.0500		1.8
3	Phenol-2	0.015	0.1004	0.1000		0.4
4	Phenol-2	0.032	0.2532	0.2500		1.3
5	Phenol-2	0.058	0.4978	0.5000		-0.4
6	Phenol-2	0.111	0.9960	1.0000		-0.4
7	Phenol-2	0.218	2.0021	2.0000		0.1

05/11/2021
AM