



### Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114391

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	05/11/2021	17:15
2	CAL2	1	Water	NA	NA	20.4	7.00	05/11/2021	17:16
3	CAL3	1	Water	NA	NA	20.4	10.03	05/11/2021	17:17
4	ICV	1	Water	NA	NA	20.3	7.02	05/11/2021	17:18
5	CCV1	1	Water	NA	NA	20.3	2.01	05/11/2021	17:20
6	M2364-02	1	Solid	20.02	20	20.7	8.76	05/11/2021	17:22
7	M2364-02DUP	1	Solid	20.03	20	20.8	8.77	05/11/2021	17:23
8	CCV2	1	Water	NA	NA	20.3	12.02	05/11/2021	17:25

88 114391

### WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M2364      WorkList ID : 148785      Department : Wet-Chemistry      Date : 05-11-2021 16:03:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/10/2021	Solid	M2364-02	Corrosivity	Cool 4 deg C	ENTA05	P23	AST2-COPR4-1	05/10/2021	9045D

Date/Time 05-11-21 16:10  
 Raw Sample Received by: JP (wet Chem)  
 Raw Sample Relinquished by: KS (Wetchem)

Date/Time 05-11-21 17:35  
 Raw Sample Received by: KS (Wetchem)  
 Raw Sample Relinquished by: JP (Wet Chem)