

LB 114394


7470A

INSTRUMENT ID: LY1

Linear

 μ Abs.:

36342



A= 0.0000e+000

B= 2.7607e-004

C= -1.5675e-002

Rho= 0.9999935

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.004	-0.004	42	0.000	42					
0.2	0.200	0.217	0.017	842	0.0 %	842					
2.5	2.500	2.491	-0.009	9081	0.0 %	9081					
5.0	5.000	4.998	-0.002	18162	0.0 %	18162					
7.5	7.500	7.480	-0.020	27153	0.0 %	27153					
10.0	10.000	10.017	0.017	36342	0.0 %	36342					

0.0000

LB114394 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 11 May 2021 17:50:36					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	42	-	0.0000	7470A	PPB	11 May 2021 17:54:34	S	Std
0.2 - 1	50.2	842	-	0.2000	7470A	PPB	11 May 2021 17:56:51	S	Std
2.5 - 1	52.5	9081	-	2.5000	7470A	PPB	11 May 2021 17:59:08	S	Std
5.0 - 1	55	18162	-	5.0000	7470A	PPB	11 May 2021 18:01:24	S	Std
7.5 - 1	57.5	27153	-	7.5000	7470A	PPB	11 May 2021 18:03:40	S	Std
10.0 - 1	59	36342	-	10.0000	7470A	PPB	11 May 2021 18:05:57	S	Std
ICV21 - 1	ICV21	14487	3.9837	-	7470A	PPB	11 May 2021 18:08:40	U	SMPL
ICB21 - 1	ICB21	42	-0.0041	-	7470A	PPB	11 May 2021 18:13:23	U	SMPL
CCV55 - 1	CCV55	18091	4.9787	-	7470A	PPB	11 May 2021 18:15:37	U	SMPL
CCB55 - 1	CCB55	-57	-0.0314	-	7470A	PPB	11 May 2021 18:17:52	U	SMPL
CRA - 1	CRA	761	0.1944	-	7470A	PPB	11 May 2021 18:20:08	U	SMPL
HighStd - 1	HighStd	35939	9.9060	-	7470A	PPB	11 May 2021 18:22:22	U	SMPL
ChkStd - 1	ChkStd	24813	6.8344	-	7470A	PPB	11 May 2021 18:24:37	U	SMPL
PB136328BL - 1	PBW	-110	-0.0460	-	7470A	PPB	11 May 2021 18:26:54	U	SMPL
PB136328BS - 1	LCSW	12766	3.5086	-	7470A	PPB	11 May 2021 18:29:11	U	SMPL
M2295-13 - 1	DRUM-B-C-E	164	0.0296	-	7470A	PPB	11 May 2021 18:31:27	U	SMPL
M2322-11 - 1	WC-20210505	140	0.0230	-	7470A	PPB	11 May 2021 18:33:43	U	SMPL
M2322-11DUP - 1	WC-20210505DUP	146	0.0246	-	7470A	PPB	11 May 2021 18:35:59	U	SMPL
M2322-11MS - 1	WC-20210505MS	12222	3.3584	-	7470A	PPB	11 May 2021 18:38:16	U	SMPL
M2322-11MSD - 1	WC-20210505MSD	12886	3.5418	-	7470A	PPB	11 May 2021 18:42:58	U	SMPL
CCV56 - 1	CCV56	18260	5.0254	-	7470A	PPB	11 May 2021 18:45:14	U	SMPL
CCB56 - 1	CCB56	-32	-0.0245	-	7470A	PPB	11 May 2021 18:47:29	U	SMPL
PB136329BL - 1	PBW	16	-0.0113	-	7470A	PPB	11 May 2021 18:49:45	U	SMPL
PB136329BS - 1	LCSW	13901	3.8220	-	7470A	PPB	11 May 2021 18:56:18	U	SMPL
M2302-01 - 1	104-SB-1	802	0.2057	-	7470A	PPB	11 May 2021 18:58:33	U	SMPL
M2302-01MS - 1	104-SB-1MS	15637	4.3012	-	7470A	PPB	11 May 2021 19:03:06	U	SMPL
M2302-01MSD - 1	104-SB-1MSD	16238	4.4671	-	7470A	PPB	11 May 2021 19:05:22	U	SMPL
M2302-01DUP - 1	104-SB-1DUP	692	0.1754	-	7470A	PPB	11 May 2021 19:10:39	U	SMPL
PB136330BL - 1	PBW	7	-0.0137	-	7470A	PPB	11 May 2021 19:13:44	U	SMPL
PB136330BS - 1	LCSW	12822	3.5241	-	7470A	PPB	11 May 2021 19:16:01	U	SMPL
M2253-05 - 1	41764	20808	5.7288	-	7470A	PPB	11 May 2021 19:18:18	U	SMPL
M2261-01 - 1	OFPOL-MW-07	-81	-0.0380	-	7470A	PPB	11 May 2021 19:20:34	U	SMPL
CCV57 - 1	CCV57	18070	4.9729	-	7470A	PPB	11 May 2021 19:22:50	U	SMPL
CCB57 - 1	CCB57	-96	-0.0422	-	7470A	PPB	11 May 2021 19:25:06	U	SMPL
M2261-02 - 1	SB-2(14-17)	220	0.0451	-	7470A	PPB	11 May 2021 19:27:21	U	SMPL
M2261-03 - 1	SB-2(24-27)	270	0.0589	-	7470A	PPB	11 May 2021 19:29:36	U	SMPL
M2261-04 - 1	OFPOL-MW-07	-10	-0.0184	-	7470A	PPB	11 May 2021 19:31:52	U	SMPL
M2261-05 - 1	SB-2(14-17)	26	-0.0085	-	7470A	PPB	11 May 2021 19:34:08	U	SMPL
M2261-06 - 1	SB-2(24-27)	37	-0.0055	-	7470A	PPB	11 May 2021 19:36:23	U	SMPL
M2295-12 - 1	DRUM-B-C-E	1046	0.2731	-	7470A	PPB	11 May 2021 19:38:40	U	SMPL
M2295-12DUP - 1	DRUM-B-C-EDUP	1014	0.2643	-	7470A	PPB	11 May 2021 19:40:56	U	SMPL
M2295-12MS - 1	DRUM-B-C-EMS	17273	4.7529	-	7470A	PPB	11 May 2021 19:43:13	U	SMPL
M2295-12MSD - 1	DRUM-B-C-EMSD	16627	4.5745	-	7470A	PPB	11 May 2021 19:45:29	U	SMPL
M2298-01 - 1	CAL-AQIDW-20210504	-83	-0.0386	-	7470A	PPB	11 May 2021 19:47:45	U	SMPL
CCV58 - 1	CCV58	18013	4.9572	-	7470A	PPB	11 May 2021 19:50:00	U	SMPL
CCB58 - 1	CCB58	-53	-0.0303	-	7470A	PPB	11 May 2021 19:52:16	U	SMPL
M2307-01 - 1	MW-05	128	0.0197	-	7470A	PPB	11 May 2021 19:54:31	U	SMPL
M2307-02 - 1	MW-18	1136	0.2979	-	7470A	PPB	11 May 2021 19:56:47	U	SMPL
M2322-10 - 1	WC-20210505	219	0.0448	-	7470A	PPB	11 May 2021 19:59:02	U	SMPL
M2322-11LX5 - 1		18	-0.0107	-	7470A	PPB	11 May 2021 20:01:18	U	SMPL
M2322-11A - 1		15894	4.3722	-	7470A	PPB	11 May 2021 20:03:34	U	SMPL
M2302-01LX5 - 1		-77	-0.0369	-	7470A	PPB	11 May 2021 20:05:50	U	SMPL
M2302-01A - 1		16454	4.5268	-	7470A	PPB	11 May 2021 20:08:07	U	SMPL
M2295-12LX5 - 1		95	0.0106	-	7470A	PPB	11 May 2021 20:10:23	U	SMPL
M2295-12A - 1		16486	4.5356	-	7470A	PPB	11 May 2021 20:12:42	U	SMPL
CCV59 - 1	CCV59	17901	4.9262	-	7470A	PPB	11 May 2021 20:14:58	U	SMPL
CCB59 - 1	CCB59	-63	-0.0331	-	7470A	PPB	11 May 2021 20:17:14	U	SMPL
PB136328TB - 1	PB136225TB	2	-0.0151	-	7470A	PPB	11 May 2021 20:19:30	U	SMPL
PB136329TB - 1	PB136190TB	23	-0.0093	-	7470A	PPB	11 May 2021 20:21:45	U	SMPL
CCV60 - 1	CCV60	18118	4.9862	-	7470A	PPB	11 May 2021 20:24:00	U	SMPL
CCB60 - 1	CCB60	-51	-0.0298	-	7470A	PPB	11 May 2021 20:26:15	U	SMPL