



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 05/13/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 05/12/2021 16:00 DATE & TIME OUT1: 05/13/2021 08:00OVEN IN TEMP2: 103 °COVEN OUT TEMP2: 103 °CDATE & TIME IN2: 05/13/2021 08:30 DATE & TIME OUT2: 05/13/2021 10:10Run Number: LB114408BalanceID: WC SC-3OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB114408BL	LB114408BL	148.6379	148.6379	100	148.6380	148.6380	148.6380	0.0001	1
2	M2364-03	AST2-COPR4-1	152.7523	152.7523	100	152.7545	152.7545	152.7545	0.0022	22
3	M2364-03DUP	AST2-COPR4-1DUP	142.7687	142.7687	100	142.7709	142.7709	142.7709	0.0022	22

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

LB 114408

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ASTM TS M2364 WorkList ID : 148833 Department : Wet-Chemistry Date : 05-12-2021 13:03:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/18/2021	Solid	M2364-03	ASTM TS	Cool 4 deg C	ENTA05	P23	AST2-COPR4-1	05/10/2021	SM2540 B

Date/Time 05/12/21 16:10
Raw Sample Received by: JP (wt) (2mm)
Raw Sample Relinquished by: JP (wt) (2mm)

Date/Time 05/12/21 16:10
Raw Sample Received by: JP (wt) (2mm)
Raw Sample Relinquished by: JP (wt) (2mm)