


LB1140114

7471B

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 2.2209e-004

C= -1.6535e-002

Rho= 0.9999854

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	gnd
0.0	0.000	0.005	0.005	99	0.000	99					
0.2	0.200	0.221	0.021	1068	0.0 %	1068					
2.5	2.500	2.481	-0.019	11248	0.0 %	11248					
5.0	5.000	4.966	-0.034	22437	0.0 %	22437					
7.5	7.500	7.521	0.021	33939	0.0 %	33939					
10.0	10.000	10.005	0.005	45125	0.0 %	45125					

gnd
11
-1
-1
0
0

LB114414 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 May 2021 14:55:34

Sample ID	Extended ID	μ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	99	-	0.0000	7471B	PPB	12 May 2021 14:56:19	S	Std
0.2 - 1	50.2	1068	-	0.2000	7471B	PPB	12 May 2021 14:58:36	S	Std
2.5 - 1	52.5	11248	-	2.5000	7471B	PPB	12 May 2021 15:00:53	S	Std
5.0 - 1	55	22437	-	5.0000	7471B	PPB	12 May 2021 15:03:09	S	Std
7.5 - 1	57.5	33939	-	7.5000	7471B	PPB	12 May 2021 15:05:26	S	Std
10.0 - 1	510	45125	-	10.0000	7471B	PPB	12 May 2021 15:07:43	S	Std
ICV24 - 1	ICV24	17696	3.9135	-	7471B	PPB	12 May 2021 15:10:30	U	SMPL
ICB24 - 1	ICB24	-93	-0.0372	-	7471B	PPB	12 May 2021 15:12:44	U	SMPL
CCV67 - 1	CCV67	22263	4.9278	-	7471B	PPB	12 May 2021 15:15:00	U	SMPL
CCB67 - 1	CCB67	-108	-0.0405	-	7471B	PPB	12 May 2021 15:17:15	U	SMPL
CRA - 1	CRA	1032	0.2127	-	7471B	PPB	12 May 2021 15:19:31	U	SMPL
HighStd - 1	HighStd	43155	9.5676	-	7471B	PPB	12 May 2021 15:21:46	U	SMPL
ChkStd - 1	ChkStd	33039	7.3210	-	7471B	PPB	12 May 2021 15:26:55	U	SMPL
PB136357BL - 1	PBS	-98	-0.0383	-	7471B	PPB	12 May 2021 15:29:10	U	SMPL
PB136357BS - 1	LCSS	17998	3.9806	-	7471B	PPB	12 May 2021 15:31:27	U	SMPL
M2354-01 - 1	EO-01-051121	5627	1.2331	-	7471B	PPB	12 May 2021 15:33:44	U	SMPL
M2356-01 - 1	OR-03-051121	3241	0.7032	-	7471B	PPB	12 May 2021 15:46:06	U	SMPL
M2356-01DUP - 1	OR-03-051121DUP	3438	0.7470	-	7471B	PPB	12 May 2021 15:48:22	U	SMPL
M2356-01MS - 1	OR-03-051121MS	22766	5.0395	-	7471B	PPB	12 May 2021 15:50:39	U	SMPL
M2356-01MSD - 1	OR-03-051121MSD	22792	5.0453	-	7471B	PPB	12 May 2021 15:52:55	U	SMPL
CCV68 - 1	CCV68	22961	5.0828	-	7471B	PPB	12 May 2021 15:57:57	U	SMPL
CCB68 - 1	CCB68	-90	-0.0365	-	7471B	PPB	12 May 2021 16:00:12	U	SMPL
M2361-01 - 1	CL-02-051121	6235	1.3682	-	7471B	PPB	12 May 2021 16:02:29	U	SMPL
M2356-01LX5 - 1		617	0.1205	-	7471B	PPB	12 May 2021 16:04:44	U	SMPL
M2356-01A - 1		22730	5.0315	-	7471B	PPB	12 May 2021 16:07:00	U	SMPL
CCV69 - 1	CCV69	24030	5.3202	-	7471B	PPB	12 May 2021 16:09:15	U	SMPL
CCB69 - 1	CCB69	-62	-0.0303	-	7471B	PPB	12 May 2021 16:11:32	U	SMPL