



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114420

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/12/2021	17:00
2	CAL2	1	Water	NA	NA	20.3	7.00	05/12/2021	17:01
3	CAL3	1	Water	NA	NA	20.3	10.02	05/12/2021	17:02
4	ICV	1	Water	NA	NA	20.3	6.99	05/12/2021	17:05
5	CCV1	1	Water	NA	NA	20.3	2.01	05/12/2021	17:06
6	M2367-02	1	Solid	20.02	20	21.5	6.46	05/12/2021	17:09
7	M2367-06	1	Solid	20.03	20	21.3	6.38	05/12/2021	17:10
8	M2370-02	1	Solid	20.04	20	21.3	6.73	05/12/2021	17:12
9	M2371-02	1	Solid	20.02	20	21.3	6.07	05/12/2021	17:15
10	M2371-02DUP	1	Solid	20.03	20	21.4	6.08	05/12/2021	17:16
11	CCV2	1	Water	NA	NA	20.4	12.02	05/12/2021	17:18

WORKLIST(Hardcopy Internal Chain)

1811420

WorkList Name : CORROSIONITY M2367 WorkList ID : 148842 Department : Wet-Chemistry Date : 05-12-2021 16:47:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/12/2021	Solid	M2367-02	Corrosivity	Cool 4 deg C	PSEG03	E21	PV-SP1	05/12/2021	9045D
05/12/2021	Solid	M2367-06	Corrosivity	Cool 4 deg C	PSEG03	E21	PV-SP2	05/12/2021	9045D
05/12/2021	Solid	M2370-02	Corrosivity	Cool 4 deg C	PSEG03	E21	367	05/12/2021	9045D
05/12/2021	Solid	M2371-02	Corrosivity	Cool 4 deg C	PSEG03	E21	VNJ-206	05/12/2021	9045D

Date/Time 05.12.21 16:50
 Raw Sample Received by: SP (Wet-Chem)
 Raw Sample Relinquished by: KS (Wet-Chem)

Date/Time 05.12.21 17:35
 Raw Sample Received by: KS (Wet-Chem)
 Raw Sample Relinquished by: SP (Wet-Chem)