



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114421

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/12/2021	17:00
2	CAL2	1	Water	NA	NA	20.3	7.00	05/12/2021	17:01
3	CAL3	1	Water	NA	NA	20.3	10.02	05/12/2021	17:02
4	ICV	1	Water	NA	NA	20.3	6.99	05/12/2021	17:05
5	CCV1	1	Water	NA	NA	20.3	2.01	05/12/2021	17:06
6	M2367-09	1	Water	NA	NA	20.5	6.18	05/12/2021	17:19
7	M2367-09DUP	1	Water	NA	NA	20.6	6.19	05/12/2021	17:20
8	M2368-01	1	Water	NA	NA	20.3	5.69	05/12/2021	17:22
9	M2368-03	1	Water	NA	NA	20.3	5.61	05/12/2021	17:23
10	M2368-04	1	Water	NA	NA	20.4	5.71	05/12/2021	17:25
11	M2368-05	1	Water	NA	NA	20.4	5.69	05/12/2021	17:26
12	M2368-06	1	Water	NA	NA	20.2	7.56	05/12/2021	17:28
13	M2372-14	1	Water	NA	NA	20.3	7.03	05/12/2021	17:29
14	CCV2	1	Water	NA	NA	20.4	12.02	05/12/2021	17:30

WORKLIST(Hardcopy Internal Chain)

28 111121

WorkList Name : ph m2371 WorkList ID : 148843 Department : Wet-Chemistry Date : 05-12-2021 16:48:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/12/2021	Water	M2367-09	pH	Cool 4 deg C	PSEG03	E21	PV-WATER	05/12/2021	9040C
05/12/2021	Water	M2368-01	pH	Cool 4 deg C	PSEG03	E21	COMP-33184-33186	05/12/2021	9040C
05/12/2021	Water	M2368-03	pH	Cool 4 deg C	PSEG03	E21	33184	05/12/2021	9040C
05/12/2021	Water	M2368-04	pH	Cool 4 deg C	PSEG03	E21	33185	05/12/2021	9040C
05/12/2021	Water	M2368-05	pH	Cool 4 deg C	PSEG03	E21	33186	05/12/2021	9040C
05/12/2021	Water	M2368-06	pH	Cool 4 deg C	PSEG03	E21	32644	05/12/2021	9040C
05/12/2021	Water	M2372-14	pH	Cool 4 deg C	PSEG03	E21	TOTE-1-2	05/12/2021	9040C

Date/Time 05.12.21 16:50
 Raw Sample Received by: JS (Wet Chem)
 Raw Sample Relinquished by: JS (Wet Chem)

Date/Time 05.12.21 17:35
 Raw Sample Received by: JS (Wet Chem)
 Raw Sample Relinquished by: JS (Wet Chem)