


LB114429

SFAM01.1

INSTRUMENT ID: CV2



A= 0.0000e+000

B= 2.4330e-004 *slope*C= -2.7467e-003 *intercept*

Rho= 0.9999450

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.014	0.014	68	0.000	68					1
0.05	0.050					204					2
0.20	0.200	0.203	0.003	844	0.0 %	844					2
2.50	2.500	2.438	-0.062	10030	0.0 %	10030					-2
5.00	5.000	5.042	0.042	20734	0.0 %	20734					1
7.50	7.500	7.542	0.042	31008	0.0 %	31008					1
10.0	10.000	9.962	-0.038	40958	0.0 %	40958					0

LB114429

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	68	-		0 SFAM01.1	PPB	5/13/2021 9:21	Std
	0.2 S01	844	-		0.2 SFAM01.1	PPB	5/13/2021 9:23	Std
	2.5 S02	10030	-		2.5 SFAM01.1	PPB	5/13/2021 9:26	Std
	5 S03	20734	-		5 SFAM01.1	PPB	5/13/2021 9:28	Std
	7.5 S04	31008	-		7.5 SFAM01.1	PPB	5/13/2021 9:30	Std
	10 S05	40958	-		10 SFAM01.1	PPB	5/13/2021 9:32	Std
ICV021	ICV021	16128	3.9213 -		SFAM01.1	PPB	5/13/2021 9:35	SMPL
ICB021	ICB021	-112	-0.03 -		SFAM01.1	PPB	5/13/2021 9:37	SMPL
CCV072	CCV072	20740	5.0434 -		SFAM01.1	PPB	5/13/2021 9:40	SMPL
CCB072	CCB072	-120	-0.0319 -		SFAM01.1	PPB	5/13/2021 9:42	SMPL
PB136372BL	PBW372	34	0.0055 -		SFAM01.1	PPB	5/13/2021 9:44	SMPL
M2378-01	MBG1Y7	102	0.0221 -		SFAM01.1	PPB	5/13/2021 9:46	SMPL
M2378-02	MBG1Y7D	115	0.0252 -		SFAM01.1	PPB	5/13/2021 9:49	SMPL
M2378-03	MBG1Y7S	4307	1.0452 -		SFAM01.1	PPB	5/13/2021 9:51	SMPL
M2378-04	MBG1Z7	48	0.0089 -		SFAM01.1	PPB	5/13/2021 9:53	SMPL
M2378-05	MBG205	56	0.0109 -		SFAM01.1	PPB	5/13/2021 9:55	SMPL
M2378-06	MBG1Y8	73	0.015 -		SFAM01.1	PPB	5/13/2021 9:58	SMPL
PB136373BL	PBW373	36	0.006 -		SFAM01.1	PPB	5/13/2021 10:00	SMPL
M2379-01	MBG208	120	0.0264 -		SFAM01.1	PPB	5/13/2021 10:02	SMPL
M2379-02	MBG208D	161	0.0364 -		SFAM01.1	PPB	5/13/2021 10:05	SMPL
M2379-03	MBG208S	4898	1.189 -		SFAM01.1	PPB	5/13/2021 10:09	SMPL
M2379-04	MBG233	-3	-0.0035 -		SFAM01.1	PPB	5/13/2021 10:11	SMPL
M2379-05	MBG209	43	0.0077 -		SFAM01.1	PPB	5/13/2021 10:14	SMPL
CCV073	CCV073	20689	5.031 -		SFAM01.1	PPB	5/13/2021 10:16	SMPL
CCB073	CCB073	-87	-0.0239 -		SFAM01.1	PPB	5/13/2021 10:18	SMPL