



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB114444

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP91069
Residual chlorine ICV-LCS, 0.4PPM	WP91068
Chlorine Calibration std, 0.1ppm	WP91063
Chlorine Calibration std, 0.2ppm	WP91064
Chlorine Calibration std, 0.8ppm	WP91066
Chlorine Calibration std, 0.0ppm	WP91062
Chlorine Calibration std, 1.6ppm	WP91067
Residual Chlorine LOQ 0.1ppm	WP91070
Residual Chlorine Calibration and CCV std, 0	WP91065
Total Chlorine Powder Pillows	W2249

Intercept: 0.0014

Slope: 1.0263

Regression: 0.999973

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	0.00		05/13/2021	13:05
2	CAL2	0.1	1	0.000	0.100	0.100	0.10	4	05/13/2021	13:09
3	CAL3	0.2	1	0.000	0.210	0.210	0.20	-1.5	05/13/2021	13:13
4	CAL4	0.4	1	0.000	0.410	0.410	0.40	0.5	05/13/2021	13:16
5	CAL5	0.8	1	0.000	0.830	0.830	0.81	-0.9	05/13/2021	13:20
6	CAL6	1.6	1	0.000	1.640	1.640	1.60	0.2	05/13/2021	13:24



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB114444

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4200	0.4200	0.4080	05/13/2021	13:28
2	ICB				1	0.0000	0.0000	0.0000	-0.0010	05/13/2021	13:31
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.3980	05/13/2021	13:35
4	CCB1				1	0.0000	0.0100	0.0100	0.0080	05/13/2021	13:39
5	LB114444BL	50	50		1	0.0000	0.0200	0.0200	0.0180	05/13/2021	13:42
6	LB114444BS	50	50	0.4	1	0.0000	0.4200	0.4200	0.4080	05/13/2021	13:46
7	M2262-07	50	50		1	0.0000	0.0600	0.0600	0.0570	05/13/2021	13:49
8	M2262-08	50	50		1	0.0000	0.1100	0.1100	0.1060	05/13/2021	13:53
9	CCV2			0.4	1	0.0000	0.4200	0.4200	0.4080	05/13/2021	13:57
10	CCB2				1	0.0000	0.0100	0.0100	0.0080	05/13/2021	14:00