

LB 114 445


SFAM01.1

INSTRUMENT ID: CV2

Linear

μ Abs.:

4185E



A= 0.0000e+000

B= 2.3932e-004 *slope*

C= 3.5027e-002 *y-intercept*

Rho= 0.9998916

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.045	0.045	43	0.000	43					
0.05	0.050					204					
0.20	0.200	0.239	0.039	854	0.0 %	854					
2.50	2.500	2.451	-0.049	10097	0.0 %	10097					
5.00	5.000	4.905	-0.095	20350	0.0 %	20350					
7.50	7.500	7.507	0.007	31222	0.0 %	31222					
10.0	10.000	10.052	0.052	41856	0.0 %	41856					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	43	-		0 SFAM01.1	PPB	5/13/2021 15:50	Std
	0.2 S01	854	-		0.2 SFAM01.1	PPB	5/13/2021 15:53	Std
	2.5 S02	10097	-		2.5 SFAM01.1	PPB	5/13/2021 15:57	Std
	5 S03	20350	-		5 SFAM01.1	PPB	5/13/2021 15:59	Std
	7.5 S04	31222	-		7.5 SFAM01.1	PPB	5/13/2021 16:01	Std
	10 S05	41856	-		10 SFAM01.1	PPB	5/13/2021 16:04	Std
ICV022	ICV022	15851	3.8284 -		SFAM01.1	PPB	5/13/2021 16:07	SMPL
ICB022	ICB022	-58	0.0211 -		SFAM01.1	PPB	5/13/2021 16:09	SMPL
CCV077	CCV077	20510	4.9434 -		SFAM01.1	PPB	5/13/2021 16:11	SMPL
CCB077	CCB077	-63	0.02 -		SFAM01.1	PPB	5/13/2021 16:13	SMPL
PB136347BL	PBW347	12	0.0379 -		SFAM01.1	PPB	5/13/2021 16:16	SMPL
M2268-01	MA0AL2	27	0.0415 -		SFAM01.1	PPB	5/13/2021 16:18	SMPL
M2268-02	MA0AS2	31	0.0424 -		SFAM01.1	PPB	5/13/2021 16:20	SMPL
M2268-03	MA0AS2D	45	0.0458 -		SFAM01.1	PPB	5/13/2021 16:22	SMPL
M2268-04	MA0AS2S	4281	1.0595 -		SFAM01.1	PPB	5/13/2021 16:25	SMPL
M2268-05	MA0AS4	7	0.0367 -		SFAM01.1	PPB	5/13/2021 16:27	SMPL
M2268-06	MA0AS6	33	0.0429 -		SFAM01.1	PPB	5/13/2021 16:29	SMPL
M2268-07	MA0AR4	49	0.0468 -		SFAM01.1	PPB	5/13/2021 16:31	SMPL
M2268-08	MA0AR6	55	0.0482 -		SFAM01.1	PPB	5/13/2021 16:34	SMPL
M2268-09	MA0AR8	46	0.046 -		SFAM01.1	PPB	5/13/2021 16:36	SMPL
M2268-10	MA0AS0	76	0.0532 -		SFAM01.1	PPB	5/13/2021 16:38	SMPL
M2268-11	MA0AL4	32	0.0427 -		SFAM01.1	PPB	5/13/2021 16:40	SMPL
M2268-12	MA0AS8	40	0.0446 -		SFAM01.1	PPB	5/13/2021 16:43	SMPL
M2268-13	MA0AT2	36	0.0436 -		SFAM01.1	PPB	5/13/2021 16:45	SMPL
M2268-14	MA0AT4	48	0.0465 -		SFAM01.1	PPB	5/13/2021 16:47	SMPL
M2268-15	MA0AX8	10270	2.4928 -		SFAM01.1	PPB	5/13/2021 16:49	SMPL
M2268-16	MA0AL6	-14	0.0317 -		SFAM01.1	PPB	5/13/2021 16:52	SMPL
M2268-17	MA0AN6	65	0.0506 -		SFAM01.1	PPB	5/13/2021 16:54	SMPL
M2268-18	MA0AN8	61	0.0496 -		SFAM01.1	PPB	5/13/2021 16:56	SMPL
M2268-19	MA0AP2	61	0.0496 -		SFAM01.1	PPB	5/13/2021 16:59	SMPL
CCV078	CCV078	20551	4.9532 -		SFAM01.1	PPB	5/13/2021 17:01	SMPL
CCB078	CCB078	-60	0.0207 -		SFAM01.1	PPB	5/13/2021 17:03	SMPL
M2268-20	MA0AY4	4	0.036 -		SFAM01.1	PPB	5/13/2021 17:05	SMPL
M2268-21	MA0AT6	48	0.0465 -		SFAM01.1	PPB	5/13/2021 17:08	SMPL
M2268-22	MA0AT8	60	0.0494 -		SFAM01.1	PPB	5/13/2021 17:10	SMPL
PB136348BL	PBW348	39	0.0444 -		SFAM01.1	PPB	5/13/2021 17:12	SMPL
M2270-01	MA0AL3	49	0.0468 -		SFAM01.1	PPB	5/13/2021 17:14	SMPL
M2270-02	MA0AS3	40	0.0446 -		SFAM01.1	PPB	5/13/2021 17:17	SMPL
M2270-03	MA0AS3D	50	0.047 -		SFAM01.1	PPB	5/13/2021 17:19	SMPL
M2270-04	MA0AS3S	4202	1.0406 -		SFAM01.1	PPB	5/13/2021 17:21	SMPL
M2270-05	MA0AS5	16	0.0389 -		SFAM01.1	PPB	5/13/2021 17:23	SMPL
M2270-06	MA0AS7	49	0.0468 -		SFAM01.1	PPB	5/13/2021 17:26	SMPL
M2270-07	MA0AR5	109	0.0611 -		SFAM01.1	PPB	5/13/2021 17:28	SMPL
M2270-08	MA0AR7	45	0.0458 -		SFAM01.1	PPB	5/13/2021 17:30	SMPL
M2270-09	MA0AR9	53	0.0477 -		SFAM01.1	PPB	5/13/2021 17:32	SMPL
M2270-10	MA0AS1	52	0.0475 -		SFAM01.1	PPB	5/13/2021 17:35	SMPL

LB114445

INSTRUMENT ID : CV1

M2270-11	MA0AL5	56	0.0484 -	SFAM01.1	PPB	5/13/2021 17:37	SMPL
M2270-12	MA0AS9	57	0.0487 -	SFAM01.1	PPB	5/13/2021 17:39	SMPL
M2270-13	MA0AT3	66	0.0508 -	SFAM01.1	PPB	5/13/2021 17:42	SMPL
M2270-14	MA0AT5	66	0.0508 -	SFAM01.1	PPB	5/13/2021 17:44	SMPL
M2270-15	MA0AX9	10063	2.4433 -	SFAM01.1	PPB	5/13/2021 17:46	SMPL
M2270-16	MA0AL7	-13	0.0319 -	SFAM01.1	PPB	5/13/2021 17:48	SMPL
CCV079	CCV079	20671	4.9819 -	SFAM01.1	PPB	5/13/2021 17:51	SMPL
CCB079	CCB079	-50	0.0231 -	SFAM01.1	PPB	5/13/2021 17:53	SMPL
M2270-17	MA0AN7	23	0.0405 -	SFAM01.1	PPB	5/13/2021 17:55	SMPL
M2270-18	MA0AN9	28	0.0417 -	SFAM01.1	PPB	5/13/2021 17:57	SMPL
M2270-19	MA0AP3	64	0.0503 -	SFAM01.1	PPB	5/13/2021 18:00	SMPL
M2270-20	MA0AY5	41	0.0448 -	SFAM01.1	PPB	5/13/2021 18:02	SMPL
M2270-21	MA0AT7	35	0.0434 -	SFAM01.1	PPB	5/13/2021 18:04	SMPL
M2270-22	MA0AT9	67	0.0511 -	SFAM01.1	PPB	5/13/2021 18:06	SMPL
CCV080	CCV080	20511	4.9436 -	SFAM01.1	PPB	5/13/2021 18:18	SMPL
CCB080	CCB080	-86	0.0144 -	SFAM01.1	PPB	5/13/2021 18:20	SMPL