



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114448

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/13/2021	16:15
2	CAL2	1	Water	NA	NA	20.2	7.00	05/13/2021	16:16
3	CAL3	1	Water	NA	NA	20.3	10.03	05/13/2021	16:17
4	ICV	1	Water	NA	NA	20.4	6.99	05/13/2021	16:18
5	CCV1	1	Water	NA	NA	20.3	2.01	05/13/2021	16:20
6	M2385-04	1	Water	NA	NA	21.7	5.07	05/13/2021	16:22
7	M2387-03	1	Water	NA	NA	20.3	5.94	05/13/2021	16:23
8	M2387-04	1	Water	NA	NA	20.2	8.25	05/13/2021	16:25
9	M2387-06	1	Water	NA	NA	20.4	5.23	05/13/2021	16:26
10	M2387-07	1	Water	NA	NA	20.1	8.69	05/13/2021	16:29
11	M2387-08	1	Water	NA	NA	20.3	6.19	05/13/2021	16:30
12	M2387-09	1	Water	NA	NA	20.3	4.59	05/13/2021	16:32
13	M2387-10	1	Water	NA	NA	20.4	4.80	05/13/2021	16:33
14	M2388-09	1	Water	NA	NA	20.8	8.04	05/13/2021	16:35
15	M2388-11	1	Water	NA	NA	20.3	9.44	05/13/2021	16:36
16	CCV2	1	Water	NA	NA	20.3	12.02	05/13/2021	16:38
17	M2388-12	1	Water	NA	NA	20.3	6.21	05/13/2021	16:40
18	M2401-02	1	Water	NA	NA	20.2	9.25	05/13/2021	16:41
19	M2401-02DUP	1	Water	NA	NA	20.3	9.26	05/13/2021	16:42
20	CCV3	1	Water	NA	NA	20.5	2.01	05/13/2021	16:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2401 W WorkList ID : 148904 Department : Wet-Chemistry Date : 05-13-2021 16:02:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/13/2021	Water	M2385-04	C pH	Cool 4 deg C	PSEG03	E21	33141	05/13/2021	9040C
05/13/2021	Water	M2387-03	C pH	Cool 4 deg C	PSEG03	E22	31055	05/13/2021	9040C
05/13/2021	Water	M2387-04	A pH	Cool 4 deg C	PSEG03	E22	41102	05/13/2021	9040C
05/13/2021	Water	M2387-06	C pH	Cool 4 deg C	PSEG03	E22	41799	05/13/2021	9040C
05/13/2021	Water	M2387-07	C pH	Cool 4 deg C	PSEG03	E22	41800	05/13/2021	9040C
05/13/2021	Water	M2387-08	C pH	Cool 4 deg C	PSEG03	E22	41801	05/13/2021	9040C
05/13/2021	Water	M2387-09	C pH	Cool 4 deg C	PSEG03	E22	30794	05/13/2021	9040C
05/13/2021	Water	M2387-10	C pH	Cool 4 deg C	PSEG03	E22	30795	05/13/2021	9040C
05/13/2021	Water	M2388-09	D pH	Cool 4 deg C	PSEG03	E22	41113	05/13/2021	9040C
05/13/2021	Water	M2388-11	D pH	Cool 4 deg C	PSEG03	E22	41101	05/13/2021	9040C
05/13/2021	Water	M2388-12	A pH	Cool 4 deg C	PSEG03	E22	41769	05/13/2021	9040C
05/13/2021	Water	M2401-02	C pH	Cool 4 deg C	ENTA04	P31	VWTP-001	05/13/2021	9040C

Date/Time 05/13/21 16:10
 Raw Sample Received by: JP (wet Chem)
 Raw Sample Relinquished by: JD CSM

Date/Time 05/13/21 17:35
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JP (wet Chem)