

LB114456


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

55786



A= 0.0000e+000

B= 1.8305e-004

C= -1.1738e-001

Rho= 0.9997674

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.046	-0.046	391	0.000	391					1
0.2	0.200	0.246	0.046	1986	0.0 %	1986					23
2.5	2.500	2.580	0.080	14737	0.0 %	14737					3
5.0	5.000	4.952	-0.048	27692	0.0 %	27692					1
7.5	7.500	7.373	-0.127	40920	0.0 %	40920					2
10.0	10.000	10.094	0.094	55786	0.0 %	55786					1

LB114456 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 May 2021 09:12:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	391	-	0.0000	7471B	PPB	14 May 2021 09:15:59	S	Std
0.2 - 1	50.2	1986	-	0.2000	7471B	PPB	14 May 2021 09:18:15	S	Std
2.5 - 1	52.5	14737	-	2.5000	7471B	PPB	14 May 2021 09:20:32	S	Std
5.0 - 1	55	27692	-	5.0000	7471B	PPB	14 May 2021 09:22:49	S	Std
7.5 - 1	57.5	40920	-	7.5000	7471B	PPB	14 May 2021 09:25:05	S	Std
10.0 - 1	510	55786	-	10.0000	7471B	PPB	14 May 2021 09:30:04	S	Std
ICV26 - 1	ICV26	20318	3.6019	-	7471B	PPB	14 May 2021 09:32:52	U	SMPL
ICB26 - 1	ICB26	269	-0.0681	-	7471B	PPB	14 May 2021 09:35:07	U	SMPL
CCV81 - 1	CCV81	27833	4.9776	-	7471B	PPB	14 May 2021 09:37:24	U	SMPL
CCB81 - 1	CCB81	259	-0.0700	-	7471B	PPB	14 May 2021 09:39:39	U	SMPL
CRA - 1	CRA	1981	0.2452	-	7471B	PPB	14 May 2021 09:41:55	U	SMPL
HighStd - 1	HighStd	56384	10.2039	-	7471B	PPB	14 May 2021 09:44:10	U	SMPL
ChkStd - 1	ChkStd	38907	7.0047	-	7471B	PPB	14 May 2021 09:46:25	U	SMPL
PB136388BL - 1	PBS	274	-0.0672	-	7471B	PPB	14 May 2021 09:51:22	U	SMPL
PB136388BS - 1	LCSS	18748	3.3145	-	7471B	PPB	14 May 2021 09:53:38	U	SMPL
M2367-01 - 1	PV-SP1	1721	0.1977	-	7471B	PPB	14 May 2021 09:55:54	U	SMPL
M2367-05 - 1	PV-SP2	1459	0.1497	-	7471B	PPB	14 May 2021 09:58:11	U	SMPL
M2369-01 - 1	RO-365	3670	0.5544	-	7471B	PPB	14 May 2021 10:00:28	U	SMPL
M2370-02 - 1	367	3754	0.5698	-	7471B	PPB	14 May 2021 10:02:45	U	SMPL
M2371-01 - 1	VNJ-206	3188	0.4662	-	7471B	PPB	14 May 2021 10:05:00	U	SMPL
CCV82 - 1	CCV82	27561	4.9278	-	7471B	PPB	14 May 2021 10:07:16	U	SMPL
CCB82 - 1	CCB82	264	-0.0691	-	7471B	PPB	14 May 2021 10:09:31	U	SMPL
M2372-01 - 1	CHRT-24702	8098	1.3650	-	7471B	PPB	14 May 2021 10:11:48	U	SMPL
M2372-03 - 1	CHRT-27051	20917	3.7116	-	7471B	PPB	14 May 2021 10:14:03	U	SMPL
M2372-05 - 1	CHRT-26598	30457	5.4579	-	7471B	PPB	14 May 2021 10:16:19	U	SMPL
M2372-07 - 1	RBR-200032	29934	5.3621	-	7471B	PPB	14 May 2021 10:18:35	U	SMPL
M2372-09 - 1	CHRT-26516	14117	2.4668	-	7471B	PPB	14 May 2021 10:20:53	U	SMPL
M2372-11 - 1	CHRT-26855	2222	0.2894	-	7471B	PPB	14 May 2021 10:23:10	U	SMPL
M2373-01 - 1	OR-02-051221	1449	0.1479	-	7471B	PPB	14 May 2021 10:25:27	U	SMPL
M2375-01 - 1	SU-03-051221	1333	0.1266	-	7471B	PPB	14 May 2021 10:27:44	U	SMPL
M2383-01 - 1	KSB-01(6-8)	644	0.0005	-	7471B	PPB	14 May 2021 10:30:00	U	SMPL
M2383-02 - 1	KSB-02(7.5-9.5)	30961	5.5501	-	7471B	PPB	14 May 2021 10:32:16	U	SMPL
CCV83 - 1	CCV83	28058	5.0187	-	7471B	PPB	14 May 2021 10:34:31	U	SMPL
CCB83 - 1	CCB83	283	-0.0656	-	7471B	PPB	14 May 2021 10:36:48	U	SMPL
M2383-03 - 1	KSB-03(7-9)	21262	3.7747	-	7471B	PPB	14 May 2021 10:43:54	U	SMPL
M2383-03DUP - 1	KSB-03(7-9)DUP	22128	3.9332	-	7471B	PPB	14 May 2021 10:46:09	U	SMPL
M2383-04 - 1	M2383-03MS	39911	7.1885	-	7471B	PPB	14 May 2021 10:54:05	U	SMPL
M2383-05 - 1	M2383-03MSD	39932	7.1923	-	7471B	PPB	14 May 2021 10:56:20	U	SMPL
M2383-06 - 1	DUP-01-05122021	13171	2.2936	-	7471B	PPB	14 May 2021 10:58:37	U	SMPL
M2383-03LX5 - 1		4649	0.7336	-	7471B	PPB	14 May 2021 11:00:55	U	SMPL
M2383-03A - 1		40290	7.2579	-	7471B	PPB	14 May 2021 11:03:11	U	SMPL
CCV84 - 1	CCV84	28073	5.0215	-	7471B	PPB	14 May 2021 11:05:28	U	SMPL
CCB84 - 1	CCB84	297	-0.0630	-	7471B	PPB	14 May 2021 11:07:46	U	SMPL