



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114460

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/13/2021	16:25
2	CAL2	1	Water	NA	NA	20.7	7.00	05/13/2021	16:26
3	CAL3	1	Water	NA	NA	20.7	10.03	05/13/2021	16:27
4	ICV	1	Water	NA	NA	20.7	7.00	05/13/2021	16:30
5	CCV1	1	Water	NA	NA	20.2	2.01	05/13/2021	16:31
6	M2392-03	1	Solid	20.02	20	21.9	11.28	05/13/2021	16:33
7	M2392-03DUP	1	Solid	20.03	20	21.8	11.29	05/13/2021	16:54
8	M2392-06	1	Solid	20.04	20	21.8	11.10	05/13/2021	17:00
9	M2392-09	1	Solid	20.02	20	21.6	10.48	05/13/2021	17:03
10	M2392-12	1	Solid	20.03	20	21.4	9.35	05/13/2021	17:05
11	M2392-15	1	Solid	20.04	20	21.5	8.98	05/13/2021	17:06
12	M2392-18	1	Solid	20.04	20	21.3	8.72	05/13/2021	17:08
13	M2392-21	1	Solid	20.02	20	21.7	10.44	05/13/2021	17:10
14	M2392-24	1	Solid	20.03	20	21.3	9.78	05/13/2021	17:12
15	CCV2	1	Water	NA	NA	20.6	12.02	05/13/2021	17:15