



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114466

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	05/14/2021	09:20
2	M2367-09		1	9	100.00	>212.0	>212.0	05/14/2021	09:45
3	M2367-09DUP		1	9	100.00	>212.0	>212.0	05/14/2021	10:10
4	M2368-01		1	8	100.00	>212.0	>212.0	05/14/2021	10:35
5	M2368-06		1	9	100.00	>212.0	>212.0	05/14/2021	10:58
6	M2372-14		1	8	100.00	>212.0	>212.0	05/14/2021	11:25
7	M2385-04		1	8	100.00	>212.0	>212.0	05/14/2021	12:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp51221 WorkList ID : 148819 Department : Wet-Chemistry Date : 05-12-2021 09:11:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/19/2021	Water	M2367-09	Flash Point	Cool 4 deg C	PSEG03	E21	PV-WATER	05/12/2021	1010A
05/19/2021	Water	M2368-01	Flash Point	Cool 4 deg C	PSEG03	E21	COMP-33184-33186	05/12/2021	1010A
05/19/2021	Water	M2368-06	Flash Point	Cool 4 deg C	PSEG03	E21	32644	05/12/2021	1010A
05/19/2021	Water	M2372-14	Flash Point	Cool 4 deg C	PSEG03	E21	TOTE-1-2	05/12/2021	1010A
05/20/2021	Water	M2385-04	Flash Point	Cool 4 deg C	PSEG03	E21	33141	05/13/2021	1010A

Date/Time 05/14/21 08:45
Raw Sample Received by: KS (watcher)
Raw Sample Relinquished by: RCCExt-Lab

Date/Time 05/14/21 12:10
Raw Sample Received by: RCCExt-Lab
Raw Sample Relinquished by: KS (watcher)