


7470A

LB114472
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36572



A= 0.0000e+000

B= 2.7279e-004

C= -1.4781e-002

Rho= 0.9999144

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	V/D
0.0	0.000	0.013	0.013	101	0.000	101					1
0.2	0.200	0.191	-0.009	756	0.0 %	756					15
2.5	2.500	2.433	-0.067	8974	0.0 %	8974					13
5.0	5.000	5.087	0.087	18704	0.0 %	18704					2
7.5	7.500	7.513	0.013	27597	0.0 %	27597					00
10.0	10.000	9.962	-0.038	36572	0.0 %	36572					0

LB114472 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 May 2021 14:22:03

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	101	-	0.0000	7470A	PPB	14 May 2021 14:24:55	S	Std
0.2 - 1	SO.2	756	-	0.2000	7470A	PPB	14 May 2021 14:27:11	S	Std
2.5 - 1	SO.5	8974	-	2.5000	7470A	PPB	14 May 2021 14:29:29	S	Std
5.0 - 1	SO.5	18704	-	5.0000	7470A	PPB	14 May 2021 14:31:46	S	Std
7.5 - 1	SO.5	27597	-	7.5000	7470A	PPB	14 May 2021 14:34:02	S	Std
10.0 - 1	SO.10	36572	-	10.0000	7470A	PPB	14 May 2021 14:36:18	S	Std
ICV27 - 1	ICV27	15103	4.1052	-	7470A	PPB	14 May 2021 14:39:07	U	SMPL
ICB27 - 1	ICB27	169	0.0313	-	7470A	PPB	14 May 2021 14:41:22	U	SMPL
CCV85 - 1	CCV85	19118	5.2004	-	7470A	PPB	14 May 2021 14:43:36	U	SMPL
CCB85 - 1	CCB85	166	0.0305	-	7470A	PPB	14 May 2021 14:45:51	U	SMPL
CRA - 1	CRA	735	0.1857	-	7470A	PPB	14 May 2021 14:48:07	U	SMPL
HighStd - 1	HighStd	36864	10.0413	-	7470A	PPB	14 May 2021 14:50:21	U	SMPL
ChkStd - 1	ChkStd	25833	7.0322	-	7470A	PPB	14 May 2021 14:52:37	U	SMPL
PB136379BL - 1	PBW	31	-0.0063	-	7470A	PPB	14 May 2021 14:57:31	U	SMPL
PB136379BS - 1	LCSW	14120	3.8370	-	7470A	PPB	14 May 2021 14:59:46	U	SMPL
M2367-02 - 1	PV-SP1	-71	-0.0341	-	7470A	PPB	14 May 2021 15:02:02	U	SMPL
M2367-06 - 1	PV-SP2	18	-0.0099	-	7470A	PPB	14 May 2021 15:04:21	U	SMPL
M2370-03 - 1	367	32	-0.0061	-	7470A	PPB	14 May 2021 15:06:37	U	SMPL
M2371-02 - 1	VNJ-206	35	-0.0052	-	7470A	PPB	14 May 2021 15:08:54	U	SMPL
M2381-01 - 1	SB-5B	476	0.1151	-	7470A	PPB	14 May 2021 15:11:10	U	SMPL
CCV86 - 1	CCV86	18903	5.1418	-	7470A	PPB	14 May 2021 15:13:24	U	SMPL
CCB86 - 1	CCB86	182	0.0349	-	7470A	PPB	14 May 2021 15:15:39	U	SMPL
M2381-01DUP - 1	SB-5BDUP	419	0.0995	-	7470A	PPB	14 May 2021 15:20:44	U	SMPL
M2381-01MS - 1	SB-5BMS	14709	3.9977	-	7470A	PPB	14 May 2021 15:23:00	U	SMPL
M2381-01MSD - 1	SB-5BMSD	15372	4.1785	-	7470A	PPB	14 May 2021 15:25:15	U	SMPL
PB136386BL - 1	PBW	-74	-0.0350	-	7470A	PPB	14 May 2021 15:27:31	U	SMPL
PB136386BS - 1	LCSW	13939	3.7876	-	7470A	PPB	14 May 2021 15:29:46	U	SMPL
M2367-09 - 1	PV-WATER	1699	0.4487	-	7470A	PPB	14 May 2021 15:32:02	U	SMPL
M2368-01 - 1	COMP-33184-33186	7958	2.1561	-	7470A	PPB	14 May 2021 15:34:19	U	SMPL
M2372-14 - 1	TOTE-1-2	1606	0.4233	-	7470A	PPB	14 May 2021 15:36:35	U	SMPL
M2383-07 - 1	KTW-01	3261	0.8748	-	7470A	PPB	14 May 2021 15:38:53	U	SMPL
M2383-08 - 1	KTW-02	3410	0.9154	-	7470A	PPB	14 May 2021 15:41:08	U	SMPL
CCV87 - 1	CCV87	18819	5.1188	-	7470A	PPB	14 May 2021 15:43:23	U	SMPL
CCB87 - 1	CCB87	217	0.0444	-	7470A	PPB	14 May 2021 15:45:39	U	SMPL
M2383-08DUP - 1	KTW-02DUP	2997	0.8028	-	7470A	PPB	14 May 2021 15:47:54	U	SMPL
M2383-09 - 1	M2383-08MS	18237	4.9601	-	7470A	PPB	14 May 2021 15:52:49	U	SMPL
M2383-10 - 1	M2383-08MSD	17824	4.8474	-	7470A	PPB	14 May 2021 15:55:04	U	SMPL
M2383-11 - 1	KTW-03	4712	1.2706	-	7470A	PPB	14 May 2021 15:57:20	U	SMPL
M2383-12 - 1	DUP-02-05112021	307	0.0690	-	7470A	PPB	14 May 2021 15:59:36	U	SMPL
M2383-13 - 1	FB-01-05112021	0	-0.0148	-	7470A	PPB	14 May 2021 16:13:18	U	SMPL
M2383-14 - 1	FB-02-05122021	4	-0.0137	-	7470A	PPB	14 May 2021 16:15:34	U	SMPL
M2387-06 - 1	41799	1058	0.2738	-	7470A	PPB	14 May 2021 16:17:51	U	SMPL
M2387-07 - 1	41800	2012	0.5341	-	7470A	PPB	14 May 2021 16:20:08	U	SMPL
M2387-08 - 1	41801	104	0.0136	-	7470A	PPB	14 May 2021 16:22:25	U	SMPL
CCV88 - 1	CCV88	18377	4.9983	-	7470A	PPB	14 May 2021 16:24:41	U	SMPL
CCB88 - 1	CCB88	-20	-0.0202	-	7470A	PPB	14 May 2021 16:26:57	U	SMPL
M2387-09 - 1	30794	640	0.1598	-	7470A	PPB	14 May 2021 16:29:13	U	SMPL
M2387-10 - 1	30795	276	0.0605	-	7470A	PPB	14 May 2021 16:31:29	U	SMPL
M2401-02 - 1	WWTP-001	602	0.1494	-	7470A	PPB	14 May 2021 16:33:44	U	SMPL
PB136379TB - 1	PB136361TB	45	-0.0025	-	7470A	PPB	14 May 2021 16:36:00	U	SMPL
M2381-01LX5 - 1		107	0.0144	-	7470A	PPB	14 May 2021 16:41:09	U	SMPL
M2381-01A - 1		14385	3.9093	-	7470A	PPB	14 May 2021 16:43:25	U	SMPL
M2383-08LX5 - 1		813	0.2070	-	7470A	PPB	14 May 2021 16:45:41	U	SMPL
M2383-08A - 1		17378	4.7258	-	7470A	PPB	14 May 2021 16:47:57	U	SMPL
CCV89 - 1	CCV89	18400	5.0045	-	7470A	PPB	14 May 2021 16:52:47	U	SMPL
CCB89 - 1	CCB89	56	0.0005	-	7470A	PPB	14 May 2021 16:55:03	U	SMPL