

LBH4481


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44588



A= 0.0000e+000

B= 2.2440e-004

C= -5.6577e-002

Rho= 0.9998162

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.020	0.020	342	0.000	342					
0.2	0.200	0.207	0.007	1176	0.0 %	1176					3
2.5	2.500	2.488	-0.012	11338	0.0 %	11338					0
5.0	5.000	4.904	-0.096	22105	0.0 %	22105					-2
7.5	7.500	7.632	0.132	34265	0.0 %	34265					2
10.0	10.000	9.949	-0.051	44588	0.0 %	44588					-1

LB114481 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 May 2021 08:28:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	342	-	0.0000	7471B	PPB	15 May 2021 08:33:41	S	Std
0.2 - 1	50.2	1176	-	0.2000	7471B	PPB	15 May 2021 08:35:58	S	Std
2.5 - 1	52.5	11338	-	2.5000	7471B	PPB	15 May 2021 08:38:15	S	Std
5.0 - 1	55	22105	-	5.0000	7471B	PPB	15 May 2021 08:40:32	S	Std
7.5 - 1	57.5	34265	-	7.5000	7471B	PPB	15 May 2021 08:42:48	S	Std
10.0 - 1	510	44588	-	10.0000	7471B	PPB	15 May 2021 08:45:06	S	Std
ICV28 - 1	ICV28	17737	3.9235	-	7471B	PPB	15 May 2021 08:47:49	U	SMPL
ICB28 - 1	ICB28	174	-0.0175	-	7471B	PPB	15 May 2021 08:50:04	U	SMPL
CCV90 - 1	CCV90	22356	4.9600	-	7471B	PPB	15 May 2021 08:52:21	U	SMPL
CCB90 - 1	CCB90	127	-0.0281	-	7471B	PPB	15 May 2021 08:54:36	U	SMPL
CRA - 1	CRA	1151	0.2017	-	7471B	PPB	15 May 2021 08:56:52	U	SMPL
HighStd - 1	HighStd	45774	10.2149	-	7471B	PPB	15 May 2021 08:59:08	U	SMPL
ChkStd - 1	ChkStd	31764	7.0712	-	7471B	PPB	15 May 2021 09:01:23	U	SMPL
PB136440BL - 1	PBS	67	-0.0415	-	7471B	PPB	15 May 2021 09:03:41	U	SMPL
PB136440BS - 1	LCSS	18882	4.1805	-	7471B	PPB	15 May 2021 09:05:59	U	SMPL
M2387-01 - 1	41789	4469	0.9463	-	7471B	PPB	15 May 2021 09:08:16	U	SMPL
M2387-02 - 1	30782	348	0.0215	-	7471B	PPB	15 May 2021 09:10:33	U	SMPL
M2388-01 - 1	41785-LIGHT	1634	0.3101	-	7471B	PPB	15 May 2021 09:12:50	U	SMPL
M2388-03 - 1	41785-DARK	1593	0.3009	-	7471B	PPB	15 May 2021 09:15:06	U	SMPL
M2388-05 - 1	41786-LIGHT	1966	0.3846	-	7471B	PPB	15 May 2021 09:17:21	U	SMPL
CCV91 - 1	CCV91	21631	4.7973	-	7471B	PPB	15 May 2021 09:20:04	U	SMPL
CCB91 - 1	CCB91	85	-0.0375	-	7471B	PPB	15 May 2021 09:32:19	U	SMPL
M2388-07 - 1	41786-DARK	2069	0.4077	-	7471B	PPB	15 May 2021 09:34:36	U	SMPL
M2389-01 - 1	BURNETT-AVE-1-3	2391	0.4800	-	7471B	PPB	15 May 2021 09:36:51	U	SMPL
M2389-03 - 1	BURNETT-AVE-4-6	1010	0.1701	-	7471B	PPB	15 May 2021 09:39:07	U	SMPL
M2389-05 - 1	CASINO-AVE-1-2	3223	0.6667	-	7471B	PPB	15 May 2021 09:41:22	U	SMPL
M2389-07 - 1	CASINO-AVE-4-6	5223	1.1154	-	7471B	PPB	15 May 2021 09:43:40	U	SMPL
M2392-02 - 1	SB-01-COMP	4373	0.9247	-	7471B	PPB	15 May 2021 09:45:56	U	SMPL
M2392-05 - 1	SB-02-COMP	15357	3.3895	-	7471B	PPB	15 May 2021 09:48:13	U	SMPL
M2392-08 - 1	SB-03-COMP	13156	2.8956	-	7471B	PPB	15 May 2021 09:50:30	U	SMPL
M2392-11 - 1	SB-04-COMP	1500	0.2800	-	7471B	PPB	15 May 2021 09:52:47	U	SMPL
M2392-11DUP - 1	SB-04-COMPDUP	1386	0.2544	-	7471B	PPB	15 May 2021 09:55:04	U	SMPL
CCV92 - 1	CCV92	21686	4.8097	-	7471B	PPB	15 May 2021 09:57:19	U	SMPL
CCB92 - 1	CCB92	235	-0.0038	-	7471B	PPB	15 May 2021 09:59:34	U	SMPL
M2392-11MS - 1	SB-04-COMPMS	20040	4.4403	-	7471B	PPB	15 May 2021 10:01:50	U	SMPL
M2392-11MSD - 1	SB-04-COMPMSD	18275	4.0443	-	7471B	PPB	15 May 2021 10:04:06	U	SMPL
M2392-14 - 1	SB-05-COMP	109365	24.4845	-	7471B	PPB	15 May 2021 10:06:22	U	SMPL
M2392-17 - 1	SB-06-COMP	120581	27.0014	-	7471B	PPB	15 May 2021 10:08:38	U	SMPL
M2392-20 - 1	SB-07-COMP	7343	1.5912	-	7471B	PPB	15 May 2021 10:10:57	U	SMPL
M2392-23 - 1	SB-08-COMP	7192	1.5573	-	7471B	PPB	15 May 2021 10:13:17	U	SMPL
M2392-11LX5 - 1		569	0.0711	-	7471B	PPB	15 May 2021 10:15:33	U	SMPL
M2392-11A - 1		19452	4.3084	-	7471B	PPB	15 May 2021 10:17:50	U	SMPL
M2392-14DLX5 - 1		22097	4.9019	-	7471B	PPB	15 May 2021 10:20:07	U	SMPL
M2392-17DLX5 - 1		23690	5.2594	-	7471B	PPB	15 May 2021 10:22:23	U	SMPL
CCV93 - 1	CCV93	21494	4.7666	-	7471B	PPB	15 May 2021 10:24:40	U	SMPL
CCB93 - 1	CCB93	226	-0.0059	-	7471B	PPB	15 May 2021 10:26:56	U	SMPL