

LB114482


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37633



A= 0.0000e+000

B= 2.6574e-004 *slope*C= -8.5021e-003 *y-intercept*

Rho= 0.9999834

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.012	0.012	79	0.000	79					1
0.05	0.050					204					1
0.20	0.200	0.214	0.014	836	0.0 %	836					7
2.50	2.500	2.462	-0.038	9298	0.0 %	9298					-2
5.00	5.000	4.991	-0.009	18814	0.0 %	18814					0
7.50	7.500	7.528	0.028	28362	0.0 %	28362					0
10.0	10.000	9.992	-0.008	37633	0.0 %	37633					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	79	-		0 SFAM01.1	PPB	5/15/2021 11:12	Std
	0.2 S01	836	-		0.2 SFAM01.1	PPB	5/15/2021 11:14	Std
	2.5 S02	9298	-		2.5 SFAM01.1	PPB	5/15/2021 11:17	Std
	5 S03	18814	-		5 SFAM01.1	PPB	5/15/2021 11:19	Std
	7.5 S04	28362	-		7.5 SFAM01.1	PPB	5/15/2021 11:21	Std
	10 S05	37633	-		10 SFAM01.1	PPB	5/15/2021 11:24	Std
ICV023	ICV023	15581	4.132 -		SFAM01.1	PPB	5/15/2021 11:26	SMPL
ICB023	ICB023	-62	-0.025 -		SFAM01.1	PPB	5/15/2021 11:29	SMPL
CCV094	CCV094	19113	5.0706 -		SFAM01.1	PPB	5/15/2021 11:31	SMPL
CCB094	CCB094	-62	-0.025 -		SFAM01.1	PPB	5/15/2021 11:33	SMPL
PB136444BL	PBW444	29	-0.0008 -		SFAM01.1	PPB	5/15/2021 11:35	SMPL
M2398-01	MBG1Y6	37	0.0013 -		SFAM01.1	PPB	5/15/2021 11:38	SMPL
M2398-02	MBG1Z0	77	0.012 -		SFAM01.1	PPB	5/15/2021 11:40	SMPL
M2398-03	MBG1Z1	45	0.0035 -		SFAM01.1	PPB	5/15/2021 11:42	SMPL
M2398-04	MBG1Z1D	50	0.0048 -		SFAM01.1	PPB	5/15/2021 11:44	SMPL
M2398-05	MBG1Z1S	4147	1.0935 -		SFAM01.1	PPB	5/15/2021 11:47	SMPL
M2398-06	MBG1Y5	370	0.0898 -		SFAM01.1	PPB	5/15/2021 11:49	SMPL
M2398-07	MBG1Y9	37	0.0013 -		SFAM01.1	PPB	5/15/2021 11:51	SMPL
M2398-08	MBG1Z2	46	0.0037 -		SFAM01.1	PPB	5/15/2021 11:53	SMPL
PB136445BL	PBW445	32	0 -		SFAM01.1	PPB	5/15/2021 11:56	SMPL
M2399-01	MBG207	82	0.0133 -		SFAM01.1	PPB	5/15/2021 11:58	SMPL
M2399-02	MBG226	66	0.009 -		SFAM01.1	PPB	5/15/2021 12:00	SMPL
M2399-03	MBG227	28	-0.0011 -		SFAM01.1	PPB	5/15/2021 12:02	SMPL
M2399-04	MBG227D	39	0.0019 -		SFAM01.1	PPB	5/15/2021 12:05	SMPL
M2399-05	MBG227S	4534	1.1964 -		SFAM01.1	PPB	5/15/2021 12:07	SMPL
M2399-06	MBG206	14	-0.0048 -		SFAM01.1	PPB	5/15/2021 12:09	SMPL
M2399-07	MBG225	35	0.0008 -		SFAM01.1	PPB	5/15/2021 12:11	SMPL
M2399-08	MBG228	48	0.0043 -		SFAM01.1	PPB	5/15/2021 12:14	SMPL
CCV095	CCV095	19479	5.1678 -		SFAM01.1	PPB	5/15/2021 12:16	SMPL
CCB095	CCB095	-60	-0.0244 -		SFAM01.1	PPB	5/15/2021 12:18	SMPL