

LB114483


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43560



A= 0.0000e+000

B= 2.2977e-004 slope

C= 1.3870e-002 $y = mx + c$

Rho= 0.9999646

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% of
0.00	0.000	0.029	0.029	65	0.000	65					=
0.05	0.050					204					13
0.20	0.200	0.225	0.025	917	0.0 %	917					-2
2.50	2.500	2.453	-0.047	10615	0.0 %	10615					-1
5.00	5.000	4.961	-0.039	21531	0.0 %	21531					0
7.50	7.500	7.510	0.010	32624	0.0 %	32624					0
10.0	10.000	10.023	0.023	43560	0.0 %	43560					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	65	-		0 SFAM01.1	PPB	5/15/2021 12:29	Std
	0.2 S01	917	-		0.2 SFAM01.1	PPB	5/15/2021 12:31	Std
	2.5 S02	10615	-		2.5 SFAM01.1	PPB	5/15/2021 12:34	Std
	5 S03	21531	-		5 SFAM01.1	PPB	5/15/2021 12:36	Std
	7.5 S04	32624	-		7.5 SFAM01.1	PPB	5/15/2021 12:38	Std
	10 S05	43560	-		10 SFAM01.1	PPB	5/15/2021 12:40	Std
ICV024	ICV024	16267	3.7516 -		SFAM01.1	PPB	5/15/2021 12:43	SMPL
ICB024	ICB024	-80	-0.0045 -		SFAM01.1	PPB	5/15/2021 12:45	SMPL
CCV096	CCV096	21360	4.9218 -		SFAM01.1	PPB	5/15/2021 12:48	SMPL
CCB096	CCB096	-38	0.0051 -		SFAM01.1	PPB	5/15/2021 12:50	SMPL
PB136441BL	PBS441	36	0.0221 -		SFAM01.1	PPB	5/15/2021 12:52	SMPL
M2247-01	MH0AB5	1393	0.3339 -		SFAM01.1	PPB	5/15/2021 12:54	SMPL
M2247-02	MH0AB6	60530	13.9219 -		SFAM01.1	PPB	5/15/2021 12:57	SMPL
M2247-03	MH0AB7	126	0.0428 -		SFAM01.1	PPB	5/15/2021 12:59	SMPL
M2247-04	MH0AB8	2341	0.5518 -		SFAM01.1	PPB	5/15/2021 13:01	SMPL
M2247-05	MH0AB9	2354	0.5548 -		SFAM01.1	PPB	5/15/2021 13:03	SMPL
M2247-06	MH0AC0	1633	0.3891 -		SFAM01.1	PPB	5/15/2021 13:06	SMPL
M2247-07	MH0AC1	983	0.2397 -		SFAM01.1	PPB	5/15/2021 13:08	SMPL
M2247-08	MH0AC2	2475	0.5826 -		SFAM01.1	PPB	5/15/2021 13:10	SMPL
M2247-09	MH0AC4	2663	0.6258 -		SFAM01.1	PPB	5/15/2021 13:13	SMPL
M2247-10	MH0AC4D	2640	0.6205 -		SFAM01.1	PPB	5/15/2021 13:15	SMPL
M2247-11	MH0AC4S	13691	3.1597 -		SFAM01.1	PPB	5/15/2021 13:17	SMPL
M2247-12	MH0AC5	809	0.1998 -		SFAM01.1	PPB	5/15/2021 13:19	SMPL
M2247-13	MH0AC6	2688	0.6315 -		SFAM01.1	PPB	5/15/2021 13:22	SMPL
M2247-14	MH0AC7	2849	0.6685 -		SFAM01.1	PPB	5/15/2021 13:24	SMPL
M2247-15	MH0AC8	697	0.174 -		SFAM01.1	PPB	5/15/2021 13:26	SMPL
M2247-16	MH0AC9	964	0.2354 -		SFAM01.1	PPB	5/15/2021 13:28	SMPL
M2247-17	MH0AD0	2222	0.5244 -		SFAM01.1	PPB	5/15/2021 13:31	SMPL
PB136442BL	PBS442	20	0.0185 -		SFAM01.1	PPB	5/15/2021 13:33	SMPL
CCV097	CCV097	21013	4.8421 -		SFAM01.1	PPB	5/15/2021 13:35	SMPL
CCB097	CCB097	-58	0.0005 -		SFAM01.1	PPB	5/15/2021 13:37	SMPL
M2289-01	MH0AM0	277	0.0775 -		SFAM01.1	PPB	5/15/2021 13:40	SMPL
M2289-02	MH0AM0D	302	0.0833 -		SFAM01.1	PPB	5/15/2021 13:42	SMPL
M2289-03	MH0AM0S	11921	2.753 -		SFAM01.1	PPB	5/15/2021 13:44	SMPL
M2289-04	MH0AM1	1484	0.3549 -		SFAM01.1	PPB	5/15/2021 13:46	SMPL
M2289-05	MH0AM2	227	0.066 -		SFAM01.1	PPB	5/15/2021 13:49	SMPL
M2289-06	MH0AM3	60202	13.8466 -		SFAM01.1	PPB	5/15/2021 13:51	SMPL
M2289-07	MH0AM4	106732	24.5378 -		SFAM01.1	PPB	5/15/2021 13:53	SMPL
M2289-08	MH0AM5	7076	1.6397 -		SFAM01.1	PPB	5/15/2021 13:56	SMPL
M2289-09	MH0AM6	9346	2.1613 -		SFAM01.1	PPB	5/15/2021 13:58	SMPL
M2289-10	MH0AU9	12019	2.7755 -		SFAM01.1	PPB	5/15/2021 14:00	SMPL
M2289-11	MH0AW0	430	0.1127 -		SFAM01.1	PPB	5/15/2021 14:02	SMPL
M2289-12	MH0AW1	522	0.1338 -		SFAM01.1	PPB	5/15/2021 14:05	SMPL
M2289-13	MH0AW2	111	0.0394 -		SFAM01.1	PPB	5/15/2021 14:07	SMPL
M2289-14	MH0AW3	192	0.058 -		SFAM01.1	PPB	5/15/2021 14:09	SMPL
M2289-15	MH0AW4	1390	0.3333 -		SFAM01.1	PPB	5/15/2021 14:11	SMPL

LB114483

INSTRUMENT ID : CV1

M2289-16	MHOAW5	964	0.2354 -	SFAM01.1	PPB	5/15/2021 14:14 SMPL
M2289-17	MHOAW6	3102	0.7266 -	SFAM01.1	PPB	5/15/2021 14:16 SMPL
M2289-18	MHOAW7	1586	0.3783 -	SFAM01.1	PPB	5/15/2021 14:18 SMPL
M2289-19	MHOAW8	6278	1.4564 -	SFAM01.1	PPB	5/15/2021 14:20 SMPL
PB136443BL	PBS443	5	0.015 -	SFAM01.1	PPB	5/15/2021 14:23 SMPL
CCV098	CCV098	22413	5.1637 -	SFAM01.1	PPB	5/15/2021 14:25 SMPL
CCB098	CCB098	-80	-0.0045 -	SFAM01.1	PPB	5/15/2021 14:27 SMPL
M2342-01	MHOAY2	211	0.0624 -	SFAM01.1	PPB	5/15/2021 14:30 SMPL
M2342-02	MHOAY2D	259	0.0734 -	SFAM01.1	PPB	5/15/2021 14:32 SMPL
M2342-03	MHOAY2S	12235	2.8251 -	SFAM01.1	PPB	5/15/2021 14:34 SMPL
M2342-04	MHOAY3	2469	0.5812 -	SFAM01.1	PPB	5/15/2021 14:36 SMPL
M2247-02DLX5	MHOAB6	12782	2.9508 -	SFAM01.1	PPB	5/15/2021 14:39 SMPL
M2289-06DLX5	MHOAM3	12199	2.8169 -	SFAM01.1	PPB	5/15/2021 14:41 SMPL
M2289-07DLX5	MHOAM4	22404	5.1617 -	SFAM01.1	PPB	5/15/2021 14:43 SMPL
CCV099	CCV099	22349	5.149 -	SFAM01.1	PPB	5/15/2021 14:45 SMPL
CCB099	CCB099	-32	0.0065 -	SFAM01.1	PPB	5/15/2021 14:48 SMPL