

LB114484


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37209



A= 0.0000e+000

B= 2.6659e-004

C= -4.6839e-002

Rho= 0.9997239

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.002	-0.002	169	0.000	169					-
0.2	0.200	0.187	-0.013	878	0.0 %	878					-7
2.5	2.500	2.451	-0.049	9371	0.0 %	9371					-2
5.0	5.000	5.031	0.031	19046	0.0 %	19046					1
7.5	7.500	7.660	0.160	28907	0.0 %	28907					2
10.0	10.000	9.873	-0.127	37209	0.0 %	37209					-1

LB114484 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 16 May 2021 09:37:15				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	169	-	0.0000	7470A	PPB	16 May 2021 09:40:02	S Std
0.2 - 1	50.2	878	-	0.2000	7470A	PPB	16 May 2021 09:42:19	S Std
2.5 - 1	52.5	9371	-	2.5000	7470A	PPB	16 May 2021 09:44:36	S Std
5.0 - 1	55	19046	-	5.0000	7470A	PPB	16 May 2021 09:46:53	S Std
7.5 - 1	57.5	28907	-	7.5000	7470A	PPB	16 May 2021 09:49:10	S Std
10.0 - 1	50	37209	-	10.0000	7470A	PPB	16 May 2021 09:58:09	S Std
ICV29 - 1	ICV29	15177	3.9993	-	7470A	PPB	16 May 2021 10:00:51	U SMPL
ICB29 - 1	ICB29	103	-0.0194	-	7470A	PPB	16 May 2021 10:03:06	U SMPL
CCV01 - 1	CCV01	19196	5.0707	-	7470A	PPB	16 May 2021 10:05:21	U SMPL
CCB01 - 1	CCB01	105	-0.0188	-	7470A	PPB	16 May 2021 10:07:36	U SMPL
CRA - 1	CRA	871	0.1854	-	7470A	PPB	16 May 2021 10:09:51	U SMPL
HighStd - 1	HighStd	37237	9.8803	-	7470A	PPB	16 May 2021 10:12:06	U SMPL
ChkStd - 1	ChkStd	26190	6.9353	-	7470A	PPB	16 May 2021 10:14:21	U SMPL
PB136416BL - 1	PBW	4	-0.0458	-	7470A	PPB	16 May 2021 10:16:39	U SMPL
PB136416BS - 1	LCSW	16513	4.3554	-	7470A	PPB	16 May 2021 10:18:55	U SMPL
M2392-03 - 1	SB-01-COMP	251	0.0201	-	7470A	PPB	16 May 2021 10:21:11	U SMPL
M2392-03DUP - 1	SB-01-COMPDUP	160	-0.0042	-	7470A	PPB	16 May 2021 10:23:27	U SMPL
M2392-03MS - 1	SB-01-COMPMS	15768	4.1568	-	7470A	PPB	16 May 2021 10:25:44	U SMPL
M2392-03MSD - 1	SB-01-COMPMSD	17190	4.5359	-	7470A	PPB	16 May 2021 10:28:00	U SMPL
M2392-06 - 1	SB-02-COMP	69	-0.0284	-	7470A	PPB	16 May 2021 10:30:15	U SMPL
CCV02 - 1	CCV02	19286	5.0947	-	7470A	PPB	16 May 2021 10:32:31	U SMPL
CCB02 - 1	CCB02	98	-0.0207	-	7470A	PPB	16 May 2021 10:34:46	U SMPL
M2392-09 - 1	SB-03-COMP	214	0.0102	-	7470A	PPB	16 May 2021 10:37:01	U SMPL
M2392-12 - 1	SB-04-COMP	203	0.0073	-	7470A	PPB	16 May 2021 10:39:17	U SMPL
M2392-15 - 1	SB-05-COMP	210	0.0091	-	7470A	PPB	16 May 2021 10:41:32	U SMPL
M2392-18 - 1	SB-06-COMP	198	0.0059	-	7470A	PPB	16 May 2021 10:43:47	U SMPL
M2392-21 - 1	SB-07-COMP	202	0.0070	-	7470A	PPB	16 May 2021 10:46:03	U SMPL
M2392-24 - 1	SB-08-COMP	195	0.0051	-	7470A	PPB	16 May 2021 10:48:19	U SMPL
PB136417BL - 1	PBW	183	0.0019	-	7470A	PPB	16 May 2021 10:50:35	U SMPL
PB136417BS - 1	LCSW	14858	3.9142	-	7470A	PPB	16 May 2021 10:52:52	U SMPL
M2385-04 - 1	33141	41	-0.0359	-	7470A	PPB	16 May 2021 10:55:09	U SMPL
M2385-04DUP - 1	33141DUP	94	-0.0218	-	7470A	PPB	16 May 2021 10:57:24	U SMPL
CCV03 - 1	CCV03	19169	5.0635	-	7470A	PPB	16 May 2021 10:59:40	U SMPL
CCB03 - 1	CCB03	119	-0.0151	-	7470A	PPB	16 May 2021 11:01:55	U SMPL
M2385-04MS - 1	33141MS	14925	3.9321	-	7470A	PPB	16 May 2021 11:06:49	U SMPL
M2385-04MSD - 1	33141MSD	14818	3.9035	-	7470A	PPB	16 May 2021 11:09:04	U SMPL
PB136416TB - 1	PB136392TB	59	-0.0311	-	7470A	PPB	16 May 2021 11:11:19	U SMPL
PB136417TB - 1	PB136408TB	150	-0.0068	-	7470A	PPB	16 May 2021 11:13:34	U SMPL
M2392-03LX5 - 1		205	0.0078	-	7470A	PPB	16 May 2021 11:15:50	U SMPL
M2392-03A - 1		15330	4.0400	-	7470A	PPB	16 May 2021 11:18:06	U SMPL
M2385-04LX5 - 1		35	-0.0375	-	7470A	PPB	16 May 2021 11:20:23	U SMPL
M2385-04A - 1		15121	3.9843	-	7470A	PPB	16 May 2021 11:22:40	U SMPL
CCV04 - 1	CCV04	18845	4.9771	-	7470A	PPB	16 May 2021 11:24:57	U SMPL
CCB04 - 1	CCB04	125	-0.0135	-	7470A	PPB	16 May 2021 11:27:12	U SMPL