


LB114485

SFAM01.1 INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5115e-004 *slope*

C= 1.5986e-002 *y-intercept*

Rho= 0.9999300

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.00	0.000	0.023	0.023	27	0.000	27					9.01
0.05	0.050					204					=
0.20	0.200	0.228	0.028	845	0.0 %	845					14
2.50	2.500	2.485	-0.015	9832	0.0 %	9832					-1
5.00	5.000	4.951	-0.049	19648	0.0 %	19648					-1
7.50	7.500	7.446	-0.054	29583	0.0 %	29583					-1
10.0	10.000	10.067	0.067	40021	0.0 %	40021					1

LB114485

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	27	-		0 SFAM01.1	PPB	5/16/2021 12:28	Std
	0.2 S01	845	-		0.2 SFAM01.1	PPB	5/16/2021 12:30	Std
	2.5 S02	9832	-		2.5 SFAM01.1	PPB	5/16/2021 12:32	Std
	5 S03	19648	-		5 SFAM01.1	PPB	5/16/2021 12:35	Std
	7.5 S04	29583	-		7.5 SFAM01.1	PPB	5/16/2021 12:40	Std
	10 S05	40021	-		10 SFAM01.1	PPB	5/16/2021 12:44	Std
ICV025	ICV025	15216	3.8375 -		SFAM01.1	PPB	5/16/2021 12:47	SMPL
ICB025	ICB025	-85	-0.0054 -		SFAM01.1	PPB	5/16/2021 12:49	SMPL
CCV005	CCV005	19621	4.9438 -		SFAM01.1	PPB	5/16/2021 12:52	SMPL
CCB005	CCB005	-128	-0.0162 -		SFAM01.1	PPB	5/16/2021 12:54	SMPL
PB136446BL	PBW446	0	0.016 -		SFAM01.1	PPB	5/16/2021 12:56	SMPL
M2425-01	MBG125	28	0.023 -		SFAM01.1	PPB	5/16/2021 12:58	SMPL
M2425-02	MBG126	40	0.026 -		SFAM01.1	PPB	5/16/2021 13:01	SMPL
M2425-03	MBG126D	45	0.0273 -		SFAM01.1	PPB	5/16/2021 13:03	SMPL
M2425-04	MBG126S	4208	1.0728 -		SFAM01.1	PPB	5/16/2021 13:05	SMPL
PB136447BL	PBW447	-23	0.0102 -		SFAM01.1	PPB	5/16/2021 13:08	SMPL
M2426-01	MBG231	25	0.0223 -		SFAM01.1	PPB	5/16/2021 13:10	SMPL
M2426-02	MBG232	30	0.0235 -		SFAM01.1	PPB	5/16/2021 13:12	SMPL
M2426-03	MBG232D	38	0.0255 -		SFAM01.1	PPB	5/16/2021 13:14	SMPL
M2426-04	MBG232S	4307	1.0977 -		SFAM01.1	PPB	5/16/2021 13:17	SMPL
CCV006	CCV006	19340	4.8733 -		SFAM01.1	PPB	5/16/2021 13:22	SMPL
CCB006	CCB006	-92	-0.0071 -		SFAM01.1	PPB	5/16/2021 13:24	SMPL