

LB114486


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4176E



A= 0.0000e+000

B= 2.4059e-004 slope

C= 1.1527e-002 y-intercept

Rho= 0.9999217

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% (1)
0.00	0.000	0.025	0.025	55	0.000	55					-
0.05	0.050					204					14
0.20	0.200	0.228	0.028	898	0.0 %	898					-2
2.50	2.500	2.450	-0.050	10137	0.0 %	10137					0
5.00	5.000	5.010	0.010	20774	0.0 %	20774					-1
7.50	7.500	7.428	-0.072	30824	0.0 %	30824					101
10.0	10.000	10.060	0.060	41766	0.0 %	41766					101

NB
5/16/21

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	55	-		0 SFAM01.1	PPB	5/16/2021 13:42	Std
	0.2 S01	898	-		0.2 SFAM01.1	PPB	5/16/2021 13:44	Std
	2.5 S02	10137	-		2.5 SFAM01.1	PPB	5/16/2021 13:46	Std
	5 S03	20774	-		5 SFAM01.1	PPB	5/16/2021 13:49	Std
	7.5 S04	30824	-		7.5 SFAM01.1	PPB	5/16/2021 13:51	Std
	10 S05	41766	-		10 SFAM01.1	PPB	5/16/2021 13:53	Std
ICV026	ICV026	16446	3.9683 -		SFAM01.1	PPB	5/16/2021 13:59	SMPL
ICB026	ICB026	-109	-0.0147 -		SFAM01.1	PPB	5/16/2021 14:01	SMPL
CCV007	CCV007	20950	5.0519 -		SFAM01.1	PPB	5/16/2021 14:03	SMPL
CCB007	CCB007	-125	-0.0185 -		SFAM01.1	PPB	5/16/2021 14:05	SMPL
PB136448BL	PBW448	52	0.024 -		SFAM01.1	PPB	5/16/2021 14:07	SMPL
M2283-01	MH0AD1	57	0.0252 -		SFAM01.1	PPB	5/16/2021 14:10	SMPL
M2283-02	MH0AD3	46	0.0226 -		SFAM01.1	PPB	5/16/2021 14:12	SMPL
M2283-03	MH0AD5	69	0.0281 -		SFAM01.1	PPB	5/16/2021 14:14	SMPL
M2283-04	MH0AD7	94	0.0341 -		SFAM01.1	PPB	5/16/2021 14:16	SMPL
M2283-05	MH0AD9	32	0.0192 -		SFAM01.1	PPB	5/16/2021 14:19	SMPL
M2283-06	MH0AD9D	187	0.0565 -		SFAM01.1	PPB	5/16/2021 14:21	SMPL
M2283-07	MH0AD9S	3847	0.9371 -		SFAM01.1	PPB	5/16/2021 14:23	SMPL
M2283-08	MH0AE1	23	0.0171 -		SFAM01.1	PPB	5/16/2021 14:26	SMPL
M2283-09	MH0AE3	63	0.0267 -		SFAM01.1	PPB	5/16/2021 14:28	SMPL
M2283-10	MH0AE5	41	0.0214 -		SFAM01.1	PPB	5/16/2021 14:30	SMPL
M2283-11	MH0AE7	39	0.0209 -		SFAM01.1	PPB	5/16/2021 14:32	SMPL
M2283-12	MH0AE9	31	0.019 -		SFAM01.1	PPB	5/16/2021 14:35	SMPL
M2283-13	MH0AF1	107	0.0373 -		SFAM01.1	PPB	5/16/2021 14:37	SMPL
M2283-14	MH0AF3	83	0.0315 -		SFAM01.1	PPB	5/16/2021 14:39	SMPL
M2283-15	MH0AF5	66	0.0274 -		SFAM01.1	PPB	5/16/2021 14:41	SMPL
M2283-16	MH0AF7	108	0.0375 -		SFAM01.1	PPB	5/16/2021 14:44	SMPL
M2283-17	MH0AF9	96	0.0346 -		SFAM01.1	PPB	5/16/2021 14:46	SMPL
M2283-18	MH0AG1	67	0.0276 -		SFAM01.1	PPB	5/16/2021 14:48	SMPL
M2283-19	MH0AG5	76	0.0298 -		SFAM01.1	PPB	5/16/2021 14:50	SMPL
CCV008	CCV008	20474	4.9374 -		SFAM01.1	PPB	5/16/2021 14:53	SMPL
CCB008	CCB008	-100	-0.0125 -		SFAM01.1	PPB	5/16/2021 14:55	SMPL
M2283-20	MH0AG7	18	0.0159 -		SFAM01.1	PPB	5/16/2021 14:57	SMPL
M2283-21	MH0AG9	37	0.0204 -		SFAM01.1	PPB	5/16/2021 15:00	SMPL
M2283-22	MH0AH1	85	0.032 -		SFAM01.1	PPB	5/16/2021 15:02	SMPL
PB136449BL	PBW449	67	0.0276 -		SFAM01.1	PPB	5/16/2021 15:04	SMPL
M2284-01	MH0AD2	41	0.0214 -		SFAM01.1	PPB	5/16/2021 15:06	SMPL
M2284-02	MH0AD4	105	0.0368 -		SFAM01.1	PPB	5/16/2021 15:09	SMPL
M2284-03	MH0AD6	43	0.0219 -		SFAM01.1	PPB	5/16/2021 15:11	SMPL
M2284-04	MH0AD8	41	0.0214 -		SFAM01.1	PPB	5/16/2021 15:13	SMPL
M2284-05	MH0AE0	49	0.0233 -		SFAM01.1	PPB	5/16/2021 15:15	SMPL
M2284-06	MH0AE0D	61	0.0262 -		SFAM01.1	PPB	5/16/2021 15:18	SMPL
M2284-07	MH0AE0S	4285	1.0425 -		SFAM01.1	PPB	5/16/2021 15:23	SMPL
M2284-08	MH0AE2	80	0.0308 -		SFAM01.1	PPB	5/16/2021 15:25	SMPL
M2284-09	MH0AE4	132	0.0433 -		SFAM01.1	PPB	5/16/2021 15:27	SMPL
M2284-10	MH0AE6	200	0.0596 -		SFAM01.1	PPB	5/16/2021 15:29	SMPL

INSTRUMENT ID : CV1

M2284-11	MHOAE8	188	0.0568 -	SFAM01.1	PPB	5/16/2021 15:32 SMPL
M2284-12	MHOAF0	58	0.0255 -	SFAM01.1	PPB	5/16/2021 15:34 SMPL
M2284-13	MHOAF2	55	0.0248 -	SFAM01.1	PPB	5/16/2021 15:36 SMPL
M2284-14	MHOAF4	82	0.0313 -	SFAM01.1	PPB	5/16/2021 15:38 SMPL
M2284-15	MHOAF6	46	0.0226 -	SFAM01.1	PPB	5/16/2021 15:41 SMPL
M2284-16	MHOAF8	123	0.0411 -	SFAM01.1	PPB	5/16/2021 15:43 SMPL
CCV009	CCV009	20241	4.8814 -	SFAM01.1	PPB	5/16/2021 15:45 SMPL
CCB009	CCB009	-100	-0.0125 -	SFAM01.1	PPB	5/16/2021 15:48 SMPL
M2284-17	MHOAG0	74	0.0293 -	SFAM01.1	PPB	5/16/2021 15:50 SMPL
M2284-18	MHOAG2	62	0.0264 -	SFAM01.1	PPB	5/16/2021 15:52 SMPL
M2284-19	MHOAG6	81	0.031 -	SFAM01.1	PPB	5/16/2021 15:54 SMPL
M2284-20	MHOAG8	66	0.0274 -	SFAM01.1	PPB	5/16/2021 15:57 SMPL
M2284-21	MHOAH0	89	0.0329 -	SFAM01.1	PPB	5/16/2021 15:59 SMPL
M2284-22	MHOAH2	77	0.0301 -	SFAM01.1	PPB	5/16/2021 16:01 SMPL
PB136450BL	PBW450	49	0.0233 -	SFAM01.1	PPB	5/16/2021 16:03 SMPL
M2285-01	MHOAH3	58	0.0255 -	SFAM01.1	PPB	5/16/2021 16:06 SMPL
M2285-02	MHOAH5	54	0.0245 -	SFAM01.1	PPB	5/16/2021 16:08 SMPL
M2285-03	MHOAH7	49	0.0233 -	SFAM01.1	PPB	5/16/2021 16:10 SMPL
M2285-04	MHOAH9	39	0.0209 -	SFAM01.1	PPB	5/16/2021 16:12 SMPL
M2285-05	MHOAJ1	45	0.0224 -	SFAM01.1	PPB	5/16/2021 16:15 SMPL
M2285-06	MHOAJ3	49	0.0233 -	SFAM01.1	PPB	5/16/2021 16:17 SMPL
M2285-07	MHOAJ5	55	0.0248 -	SFAM01.1	PPB	5/16/2021 16:19 SMPL
M2285-08	MHOAJ7	62	0.0264 -	SFAM01.1	PPB	5/16/2021 16:22 SMPL
M2285-09	MHOAP0	58	0.0255 -	SFAM01.1	PPB	5/16/2021 16:24 SMPL
M2285-10	MHOAP2	55	0.0248 -	SFAM01.1	PPB	5/16/2021 16:26 SMPL
M2285-11	MHOAQ3	56	0.025 -	SFAM01.1	PPB	5/16/2021 16:28 SMPL
M2285-12	MHOAQ4	20	0.0163 -	SFAM01.1	PPB	5/16/2021 16:31 SMPL
M2285-13	MHOAQ5	152	0.0481 -	SFAM01.1	PPB	5/16/2021 16:33 SMPL
CCV010	CCV010	20085	4.8438 -	SFAM01.1	PPB	5/16/2021 16:35 SMPL
CCB010	CCB010	-101	-0.0128 -	SFAM01.1	PPB	5/16/2021 16:37 SMPL
M2285-14	MHOAQ6	41	0.0214 -	SFAM01.1	PPB	5/16/2021 16:40 SMPL
M2285-15	MHOAQ7	399	0.1075 -	SFAM01.1	PPB	5/16/2021 16:42 SMPL
M2285-16	MHOAQ8	185	0.056 -	SFAM01.1	PPB	5/16/2021 16:44 SMPL
M2285-17	MHOAP4	213	0.0628 -	SFAM01.1	PPB	5/16/2021 16:46 SMPL
M2285-18	MHOAP6	149	0.0474 -	SFAM01.1	PPB	5/16/2021 16:49 SMPL
M2285-19	MHOAP8	445	0.1186 -	SFAM01.1	PPB	5/16/2021 16:51 SMPL
M2285-20	MHOAP8D	404	0.1087 -	SFAM01.1	PPB	5/16/2021 16:53 SMPL
M2285-21	MHOAP8S	4985	1.2109 -	SFAM01.1	PPB	5/16/2021 17:01 SMPL
M2285-22	MHOAQ0	41	0.0214 -	SFAM01.1	PPB	5/16/2021 17:03 SMPL
CCV011	CCV011	19447	4.6903 -	SFAM01.1	PPB	5/16/2021 17:05 SMPL
CCB011	CCB011	-97	-0.0118 -	SFAM01.1	PPB	5/16/2021 17:08 SMPL