



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114492

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	05/14/2021	17:30
2	CAL2	1	Water	NA	NA	20.7	7.01	05/14/2021	17:31
3	CAL3	1	Water	NA	NA	20.7	10.03	05/14/2021	17:32
4	ICV	1	Water	NA	NA	20.6	7.00	05/14/2021	17:33
5	CCV1	1	Water	NA	NA	20.6	2.01	05/14/2021	17:35
6	M2405-05	1	Water	NA	NA	23.6	3.70	05/14/2021	17:41
7	M2405-05DUP	1	Water	NA	NA	23.7	3.71	05/14/2021	17:42
8	CCV2	1	Water	NA	NA	20.6	12.02	05/14/2021	17:43

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2405 WorkList ID : 148952 Department : Wet-Chemistry Date : 05-14-2021 17:21:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/14/2021	Water	M2405-05	pH	Cool 4 deg C	PSEG03	E31	BUR-1022	05/14/2021	9040C

17:25
Date/Time 05.14.21 17:25
Raw Sample Received by: Jo (wet chem)
Raw Sample Relinquished by: KS (wet chem)

11442
Date/Time 05.14.21 17:40
Raw Sample Received by: KS (wet chem)
Raw Sample Relinquished by: Jo (wet chem)